

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032065.D **Time Analyzed:** 17:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.280	9.180	9.380	736.740	712.500	3.4
2,4-D	8.412	8.312	8.512	744.040	705.000	5.5
2,4-DCAA	7.281	7.182	7.382	774.110	750.000	3.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032065.D **Time Analyzed:** 17:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.893	9.792	9.992	722.220	712.500	1.4
2,4-D	9.003	8.902	9.102	713.740	705.000	1.2
2,4-DCAA	7.745	7.644	7.844	785.150	750.000	4.7