

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS101325

Review By	Abdul	Review On	10/14/2025 8:45:14 AM
Supervise By	mohammad	Supervise On	10/15/2025 1:15:14 AM
SubDirectory	PS101325	HP Acquire Method	HP Processing Method ps101325 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24923
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031981.D	13 Oct 2025 09:35		AR\AJ	Ok
2	I.BLK	I.BLK	PS031982.D	13 Oct 2025 10:29		AR\AJ	Ok,M
3	HSTDICC200	HSTDICC200	PS031983.D	13 Oct 2025 10:53		AR\AJ	Ok,M
4	HSTDICC500	HSTDICC500	PS031984.D	13 Oct 2025 11:17		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS031985.D	13 Oct 2025 11:41	Method fail for comp@10	AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS031986.D	13 Oct 2025 12:05		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS031987.D	13 Oct 2025 12:29		AR\AJ	Ok,M
8	HSTDICV750	ICVPS101325	PS031988.D	13 Oct 2025 12:52		AR\AJ	Ok
9	I.BLK	I.BLK	PS031989.D	13 Oct 2025 13:16		AR\AJ	Ok,M
10	HSTDCCC750	HSTDCCC750	PS031990.D	13 Oct 2025 13:41		AR\AJ	Ok
11	Q3297-01	TP1	PS031991.D	13 Oct 2025 14:05		AR\AJ	Ok,M
12	Q3297-11	TP2	PS031992.D	13 Oct 2025 14:29		AR\AJ	Ok
13	PB170051BL	PB170051BL	PS031993.D	13 Oct 2025 14:53		AR\AJ	Ok
14	PB170051BS	PB170051BS	PS031994.D	13 Oct 2025 15:17	Comp#3,9,10,15 recovery fail	AR\AJ	Not Ok
15	Q3297-01MS	TP1MS	PS031995.D	13 Oct 2025 15:41	some compound recovery fail	AR\AJ	Ok,M
16	Q3297-01MSD	TP1MSD	PS031996.D	13 Oct 2025 16:06	some compound recovery fail	AR\AJ	Ok,M
17	Q3305-01	AU-06-100725	PS031997.D	13 Oct 2025 16:30		AR\AJ	Ok
18	I.BLK	I.BLK	PS031998.D	13 Oct 2025 16:54		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS101325

Review By	Abdul	Review On	10/14/2025 8:45:14 AM
Supervise By	mohammad	Supervise On	10/15/2025 1:15:14 AM
SubDirectory	PS101325	HP Acquire Method	HP Processing Method ps101325 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC Internal Standard/PEM	PP24559
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24923

19	HSTDCCC750	HSTDCCC750	PS031999.D	13 Oct 2025 17:18		AR/AJ	Ok,M
----	------------	------------	------------	-------------------	--	-------	------

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS101425

Review By	Abdul	Review On	10/15/2025 4:09:25 PM
Supervise By	mohammad	Supervise On	10/24/2025 2:04:16 AM
SubDirectory	PS101425	HP Acquire Method	HP Processing Method ps101325 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24923
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS032000.D	14 Oct 2025 13:03		AR\AJ	Ok
2	I.BLK	I.BLK	PS032001.D	14 Oct 2025 13:27		AR\AJ	Ok,M
3	HSTDCCC750	HSTDCCC750	PS032002.D	14 Oct 2025 16:19		AR\AJ	Ok,M
4	PB170051BS	PB170051BS	PS032003.D	14 Oct 2025 17:07		AR\AJ	Ok,M
5	PB170083BL	PB170083BL	PS032004.D	14 Oct 2025 17:31		AR\AJ	Ok
6	PB170083BS	PB170083BS	PS032005.D	14 Oct 2025 17:55		AR\AJ	Ok,M
7	Q3326-01	TP-1	PS032006.D	14 Oct 2025 18:19		AR\AJ	Ok,M
8	Q3326-01MS	TP-1MS	PS032007.D	14 Oct 2025 18:44		AR\AJ	Not Ok
9	Q3326-01MSD	TP-1MSD	PS032008.D	14 Oct 2025 19:08	F Flag	AR\AJ	Not Ok
10	Q3326-05	TP-2	PS032009.D	14 Oct 2025 19:32		AR\AJ	Ok,M
11	Q3327-01	HADLEY-ROAD-TP	PS032010.D	14 Oct 2025 19:56	Surroage fail 1st column	AR\AJ	ReRun
12	Q3337-01	LAYDOWN-YARD-1	PS032011.D	14 Oct 2025 20:20	Surroage fail 1st column, F Flag	AR\AJ	Not Ok
13	Q3337-03	LAYDOWN-YARD-2	PS032012.D	14 Oct 2025 20:44	Surroage fail, F Flag	AR\AJ	Not Ok
14	I.BLK	I.BLK	PS032013.D	14 Oct 2025 21:08		AR\AJ	Ok
15	HSTDCCC750	HSTDCCC750	PS032014.D	14 Oct 2025 21:32		AR\AJ	Ok,M
16	PB170090BL	PB170090BL	PS032015.D	14 Oct 2025 22:20		AR\AJ	Ok
17	PB170090BS	PB170090BS	PS032016.D	14 Oct 2025 22:44		AR\AJ	Ok,M
18	PB170062TB	PB170062TB	PS032017.D	14 Oct 2025 23:08	F Flag	AR\AJ	Not Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS101425

Review By	Abdul	Review On	10/15/2025 4:09:25 PM
Supervise By	mohammad	Supervise On	10/24/2025 2:04:16 AM
SubDirectory	PS101425	HP Acquire Method	HP Processing Method ps101325 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24923		

19	Q3321-02	P001-DS-01	PS032018.D	14 Oct 2025 23:33	Surroage fail, F Flag	AR\AJ	ReRun
20	Q3321-02MS	P001-DS-01MS	PS032019.D	14 Oct 2025 23:57	Surroage fail, F Flag	AR\AJ	Ok,M
21	Q3321-02MSD	P001-DS-01MSD	PS032020.D	15 Oct 2025 00:21	Surroage fail, F Flag	AR\AJ	Ok,M
22	I.BLK	I.BLK	PS032021.D	15 Oct 2025 00:45		AR\AJ	Ok,M
23	HSTDCCC750	HSTDCCC750	PS032022.D	15 Oct 2025 01:09		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS101625

Review By	Abdul	Review On	10/17/2025 8:38:57 AM
Supervise By	Sohil	Supervise On	10/17/2025 1:14:38 PM
SubDirectory	PS101625	HP Acquire Method	HP Processing Method ps101325 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24923
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS032053.D	16 Oct 2025 09:11		AR\AJ	Ok
2	I.BLK	I.BLK	PS032054.D	16 Oct 2025 09:35		AR\AJ	Ok,M
3	HSTDCCC750	HSTDCCC750	PS032055.D	16 Oct 2025 09:59		AR\AJ	Ok
4	Q3345-03RE	2ND-STREET-SOILRE	PS032056.D	16 Oct 2025 10:23	Surrogate low in 1st column	AR\AJ	Confirms
5	Q3345-05RE	M-R-SOILRE	PS032057.D	16 Oct 2025 10:47	Surrogate low in 1st column	AR\AJ	Confirms
6	Q3321-02RE	P001-DS-01RE	PS032058.D	16 Oct 2025 13:01	Surroage fail, F Flag	AR\AJ	Confirms
7	Q3321-02MS	P001-DS-01MS	PS032059.D	16 Oct 2025 14:00	F flag for surrogate & comp # 1, Surrogate fail in both column	AR\AJ	Not Ok
8	Q3321-02MSD	P001-DS-01MSD	PS032060.D	16 Oct 2025 15:39	surrogate fail, Recovery & RPD fail for some compounds	AR\AJ	Not Ok
9	Q3326-01MS	TP-1MS	PS032061.D	16 Oct 2025 16:15	recovery fail for some compounds, some compounds not detected	AR\AJ	Not Ok
10	Q3326-01MSD	TP-1MSD	PS032062.D	16 Oct 2025 16:40	Surrogate fail, recovery fail for some compounds, some compounds not detected	AR\AJ	Not Ok
11	PB170062TB	PB170062TB	PS032063.D	16 Oct 2025 17:04		AR\AJ	Ok
12	I.BLK	I.BLK	PS032064.D	16 Oct 2025 17:29		AR\AJ	Ok,M
13	HSTDCCC750	HSTDCCC750	PS032065.D	16 Oct 2025 17:53		AR\AJ	Ok,M
14	PB170134BL	PB170134BL	PS032066.D	16 Oct 2025 18:17		AR\AJ	Ok,M
15	PB170134BS	PB170134BS	PS032067.D	16 Oct 2025 18:41		AR\AJ	Ok,M
16	PB170081TB	PB170081TB	PS032068.D	16 Oct 2025 19:05		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS101625

Review By	Abdul	Review On	10/17/2025 8:38:57 AM
Supervise By	Sohil	Supervise On	10/17/2025 1:14:38 PM
SubDirectory	PS101625	HP Acquire Method	HP Processing Method ps101325 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24923		

17	Q3343-04	SOIL-ROLLOFF-WC-1	PS032069.D	16 Oct 2025 19:30		AR\AJ	Ok,M
18	Q3343-04MS	SOIL-ROLLOFF-WC-1	PS032070.D	16 Oct 2025 19:54	recovery fail for some compounds	AR\AJ	Ok,M
19	Q3343-04MSD	SOIL-ROLLOFF-WC-1	PS032071.D	16 Oct 2025 20:19	recovery fail for some compounds & RPD fail for comp # 1	AR\AJ	Ok,M
20	I.BLK	I.BLK	PS032072.D	16 Oct 2025 20:43		AR\AJ	Ok
21	HSTDCCC750	HSTDCCC750	PS032073.D	16 Oct 2025 21:08		AR\AJ	Ok,M

M : Manual Integration