

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3321</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Weston Solutions, Inc.</u>	DataFile :	<u>PS032019.D</u>

		Sample				Rec		RPD		Limits			
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD	
Lab Sample ID:	Q3321-02MS	Client Sample ID:	P001-DS-01MS										
	(Column 1)												
	2,4-D		50	0	0	ug/L	0	*		65	135		
	2,4,5-TP(Silvex)	50	0	24.8	ug/L	50	*			62	139		
Lab Sample ID:	Q3321-02MS	Client Sample ID:	P001-DS-01MS										
	(Column 2)												
	2,4-D		50	0	0	ug/L	0	*		65	135		
	2,4,5-TP(Silvex)	50	0	21.4	ug/L	43	*			62	139		