

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/14/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090584.D **Time Analyzed:** 14:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.066	8.968	9.168	48.260	50.000	-3.5
Endrin	6.568	6.469	6.669	52.810	50.000	5.6
gamma-BHC (Lindane)	4.325	4.226	4.426	54.580	50.000	9.2
Heptachlor	4.924	4.825	5.025	44.400	50.000	-11.2
Heptachlor epoxide	5.685	5.586	5.786	52.400	50.000	4.8
Methoxychlor	7.488	7.390	7.590	42.200	50.000	-15.6
Tetrachloro-m-xylene	3.545	3.446	3.646	51.250	50.000	2.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/14/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090584.D **Time Analyzed:** 14:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.061	7.964	8.164	46.220	50.000	-7.6
Endrin	5.781	5.682	5.882	54.390	50.000	8.8
gamma-BHC (Lindane)	3.723	3.624	3.824	55.500	50.000	11.0
Heptachlor	4.075	3.976	4.176	50.440	50.000	0.9
Heptachlor epoxide	4.865	4.766	4.966	54.380	50.000	8.8
Methoxychlor	6.745	6.646	6.846	44.530	50.000	-10.9
Tetrachloro-m-xylene	2.875	2.776	2.976	53.770	50.000	7.5