

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/14/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090592.D **Time Analyzed:** 18:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.064	8.968	9.168	49.240	50.000	-1.5
Endrin	6.569	6.469	6.669	53.600	50.000	7.2
gamma-BHC (Lindane)	4.325	4.226	4.426	54.760	50.000	9.5
Heptachlor	4.923	4.825	5.025	44.450	50.000	-11.1
Heptachlor epoxide	5.685	5.586	5.786	52.720	50.000	5.4
Methoxychlor	7.488	7.390	7.590	42.720	50.000	-14.6
Tetrachloro-m-xylene	3.545	3.446	3.646	51.380	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/14/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090592.D **Time Analyzed:** 18:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.061	7.964	8.164	48.290	50.000	-3.4
Endrin	5.781	5.682	5.882	55.580	50.000	11.2
gamma-BHC (Lindane)	3.723	3.624	3.824	56.100	50.000	12.2
Heptachlor	4.075	3.976	4.176	51.200	50.000	2.4
Heptachlor epoxide	4.864	4.766	4.966	55.340	50.000	10.7
Methoxychlor	6.745	6.646	6.846	45.280	50.000	-9.4
Tetrachloro-m-xylene	2.875	2.776	2.976	54.340	50.000	8.7