

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090596.D **Time Analyzed:** 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.065	8.968	9.168	51.590	50.000	3.2
Endrin	6.568	6.469	6.669	55.850	50.000	11.7
gamma-BHC (Lindane)	4.325	4.226	4.426	57.040	50.000	14.1
Heptachlor	4.923	4.825	5.025	47.310	50.000	-5.4
Heptachlor epoxide	5.684	5.586	5.786	55.520	50.000	11.0
Methoxychlor	7.488	7.390	7.590	44.870	50.000	-10.3
Tetrachloro-m-xylene	3.545	3.446	3.646	53.290	50.000	6.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: ROYF02
Lab Code: ACE
SDG NO.: Q3321
GC Column: ZB-MR2
ID: 0.32 (mm)
 Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No.: CCAL03
Date Analyzed: 10/15/2025
Lab Sample No.: PSTDCCC050
Data File : PD090596.D
Time Analyzed: 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.061	7.964	8.164	51.230	50.000	2.5
Endrin	5.780	5.682	5.882	57.980	50.000	16.0
gamma-BHC (Lindane)	3.722	3.624	3.824	58.870	50.000	17.7
Heptachlor	4.074	3.976	4.176	54.440	50.000	8.9
Heptachlor epoxide	4.864	4.766	4.966	58.260	50.000	16.5
Methoxychlor	6.744	6.646	6.846	48.340	50.000	-3.3
Tetrachloro-m-xylene	2.875	2.776	2.976	57.100	50.000	14.2