

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3321</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Weston Solutions, Inc.</u>	DataFile :	<u>PD090590.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec	RPD		Limits		
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3321-02MSD (Column 1)	Client Sample ID:		P001-DS-01MSD								
	gamma-BHC (Lindane)	5	0	4.90	ug/L	98		2		60	152	20
	Heptachlor	5	0	4.00	ug/L	80		2		56	147	20
	Heptachlor epoxide	5	0	4.70	ug/L	94		2		77	143	20
	Endrin	5	0	4.70	ug/L	94		2		76	144	20
	Methoxychlor	5	0	3.90	ug/L	78		3		70	142	20
Lab Sample ID:	Q3321-02MSD (Column 2)	Client Sample ID:		P001-DS-01MSD								
	gamma-BHC (Lindane)	5	0	4.90	ug/L	98		2		60	152	20
	Heptachlor	5	0	4.40	ug/L	88		2		56	147	20
	Heptachlor epoxide	5	0	4.90	ug/L	98		2		77	143	20
	Endrin	5	0	4.90	ug/L	98		0		76	144	20
	Methoxychlor	5	0	4.00	ug/L	80		2		70	142	20