

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3321

Analytical Method: 8081B

Client: Weston Solutions, Inc.

Datafile : PD090597.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB170091BS (Column 1)	gamma-BHC (Lindane)	0.5	0.55	ug/L	109				82	129		
	Heptachlor	0.5	0.47	ug/L	94				79	127		
	Heptachlor epoxide	0.5	0.54	ug/L	107				81	124		
	Endrin	0.5	0.54	ug/L	108				81	128		
	Methoxychlor	0.5	0.41	ug/L	82				78	108		
PB170091BS (Column 2)	gamma-BHC (Lindane)	0.5	0.56	ug/L	111				82	129		
	Heptachlor	0.5	0.50	ug/L	99				79	127		
	Heptachlor epoxide	0.5	0.55	ug/L	111				81	124		
	Endrin	0.5	0.56	ug/L	112				81	128		
	Methoxychlor	0.5	0.43	ug/L	87				78	108		