

Surrogate Summary

SW-846

SDG No.: Q3321

Client: Weston Solutions, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170062TB	PB170062TB	2-Fluorophenol	150	147	98		23	138
		Phenol-d6	150	140	94		10	134
		Nitrobenzene-d5	100	90.7	91		67	132
		2-Fluorobiphenyl	100	89.4	89		52	132
		2,4,6-Tribromophenol	150	148	99		44	137
		Terphenyl-d14	100	95.7	96		42	152
PB170080BL	PB170080BL	2-Fluorophenol	150	148	99		23	138
		Phenol-d6	150	140	93		10	134
		Nitrobenzene-d5	100	95.3	95		67	132
		2-Fluorobiphenyl	100	93.2	93		52	132
		2,4,6-Tribromophenol	150	152	102		44	137
		Terphenyl-d14	100	96.8	97		42	152
PB170080BS	PB170080BS	2-Fluorophenol	150	144	96		23	138
		Phenol-d6	150	139	93		10	134
		Nitrobenzene-d5	100	86.8	87		67	132
		2-Fluorobiphenyl	100	87.0	87		52	132
		2,4,6-Tribromophenol	150	136	91		44	137
		Terphenyl-d14	100	88.1	88		42	152
Q3321-02	P001-DS-01	2-Fluorophenol	150	157	104		23	138
		Phenol-d6	150	136	91		10	134
		Nitrobenzene-d5	100	108	108		67	132
		2-Fluorobiphenyl	100	107	107		52	132
		2,4,6-Tribromophenol	150	165	110		44	137
		Terphenyl-d14	100	102	102		42	152
Q3321-02MS	P001-DS-01MS	2-Fluorophenol	150	144	96		23	138
		Phenol-d6	150	130	87		10	134
		Nitrobenzene-d5	100	102	102		67	132
		2-Fluorobiphenyl	100	102	102		52	132
		2,4,6-Tribromophenol	150	160	107		44	137
		Terphenyl-d14	100	97.2	97		42	152
Q3321-02MSD	P001-DS-01MSD	2-Fluorophenol	150	142	95		23	138
		Phenol-d6	150	131	87		10	134
		Nitrobenzene-d5	100	96.7	97		67	132
		2-Fluorobiphenyl	100	96.5	96		52	132
		2,4,6-Tribromophenol	150	155	103		44	137
		Terphenyl-d14	100	98.1	98		42	152