

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3321

Analytical Method: 8270E

Client: Weston Solutions, Inc.

DataFile: BG064500.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170080BS	Pyridine	50	37.4	ug/L	75				29	97	
	1,4-Dichlorobenzene	50	49.5	ug/L	99				76	103	
	2-Methylphenol	50	50.1	ug/L	100				69	109	
	3+4-Methylphenols	50	48.0	ug/L	96				67	106	
	Hexachloroethane	50	47.2	ug/L	94				76	118	
	Nitrobenzene	50	46.2	ug/L	92				58	106	
	Hexachlorobutadiene	50	46.0	ug/L	92				69	101	
	2,4,6-Trichlorophenol	50	51.9	ug/L	104				61	110	
	2,4,5-Trichlorophenol	50	52.2	ug/L	104				70	106	
	2,4-Dinitrotoluene	50	49.2	ug/L	98				60	115	
	Hexachlorobenzene	50	47.4	ug/L	95				73	106	
	Pentachlorophenol	100	110	ug/L	110				47	114	