

Report of Analysis

Client: Weston Solutions, Inc.
Project: RFP 918
Client Sample ID: VX1013WBSD01
Lab Sample ID: VX1013WBSD01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3321
Matrix: TCLP
% Solid: 0
Test: TCLP VOA

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-01-4	Vinyl Chloride	20.3		1	0.26	1.00	ug/L	10/13/25 11:24	VX101325
75-35-4	1,1-Dichloroethene	19.8		1	0.23	1.00	ug/L	10/13/25 11:24	VX101325
78-93-3	2-Butanone	83.2		1	0.98	5.00	ug/L	10/13/25 11:24	VX101325
56-23-5	Carbon Tetrachloride	21.5		1	0.25	1.00	ug/L	10/13/25 11:24	VX101325
67-66-3	Chloroform	19.7		1	0.25	1.00	ug/L	10/13/25 11:24	VX101325
71-43-2	Benzene	20.5		1	0.15	1.00	ug/L	10/13/25 11:24	VX101325
107-06-2	1,2-Dichloroethane	20.0		1	0.22	1.00	ug/L	10/13/25 11:24	VX101325
79-01-6	Trichloroethene	21.0		1	0.090	1.00	ug/L	10/13/25 11:24	VX101325
127-18-4	Tetrachloroethene	21.2		1	0.23	1.00	ug/L	10/13/25 11:24	VX101325
108-90-7	Chlorobenzene	20.0		1	0.12	1.00	ug/L	10/13/25 11:24	VX101325
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	49.6			74 - 12	99%	SPK: 50		
1868-53-7	Dibromofluoromethane	55.8			75 - 12	112%	SPK: 50		
2037-26-5	Toluene-d8	53.9			86 - 11	108%	SPK: 50		
460-00-4	4-Bromofluorobenzene	55.0			77 - 12	110%	SPK: 50		
INTERNAL STANDARDS									
		Area							
363-72-4	Pentafluorobenzene	139000							
540-36-3	1,4-Difluorobenzene	234000							
3114-55-4	Chlorobenzene-d5	213000							
3855-82-1	1,4-Dichlorobenzene-d4	110000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products