

Surrogate Summary

SDG No.: Q3321

Client: Weston Solutions, Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3321-02	P001-DS-01	1,2-Dichloroethane-d4	50	56.5	113		74	125
		Dibromofluoromethane	50	51.5	103		75	124
		Toluene-d8	50	50.6	101		86	113
		4-Bromofluorobenzene	50	57.3	115		77	121

Surrogate Summary

SDG No.: Q3321

Client: Weston Solutions, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
VX1013WBL01	VX1013WBL01	1,2-Dichloroethane-d4	50	49.8	100		74	125
		Dibromofluoromethane	50	48.9	98		75	124
		Toluene-d8	50	48.3	97		86	113
		4-Bromofluorobenzene	50	54.5	109		77	121
VX1013WBS01	VX1013WBS01	1,2-Dichloroethane-d4	50	50.5	101		74	125
		Dibromofluoromethane	50	55.6	111		75	124
		Toluene-d8	50	53.9	108		86	113
		4-Bromofluorobenzene	50	54.8	110		77	121
VX1013WBSD0	VX1013WBSD01	1,2-Dichloroethane-d4	50	49.6	99		74	125
		Dibromofluoromethane	50	55.8	112		75	124
		Toluene-d8	50	53.9	108		86	113
		4-Bromofluorobenzene	50	55.0	110		77	121