

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3321</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Weston Solutions, Inc.</u>	Datafile :	<u>VX048131.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1013WBSD01	Vinyl chloride	20	20.3	ug/L	102	2		65	117	19
	1,1-Dichloroethene	20	19.8	ug/L	99	2		74	110	20
	2-Butanone	100	83.2	ug/L	83	6		65	122	26
	Carbon Tetrachloride	20	21.5	ug/L	108	8		77	113	15
	Chloroform	20	19.7	ug/L	99	3		79	113	20
	Benzene	20	20.5	ug/L	103	4		82	109	15
	1,2-Dichloroethane	20	20.0	ug/L	100	4		80	115	20
	Trichloroethene	20	21.0	ug/L	105	4		77	113	15
	Tetrachloroethene	20	21.2	ug/L	106	4		67	123	15
	Chlorobenzene	20	20.0	ug/L	100	2		82	109	15