

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ROYF02
 Lab Code: ACE SDG NO.: Q3321
 Lab File ID: VX048126.D Date Analyzed: 10/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:55
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	109642	12				
UPPER LIMIT	219284	12.5				
LOWER LIMIT	54821	11.5				
EPA SAMPLE NO.						
P001-DS-01	140622	12.01				
VX1013WBL01	171429	12.01				
VX1013WBS01	119033	12.00				
VX1013WBSD01	110079	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.