

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3321

Lab File ID: VX048125.D

BFB Injection Date: 10/13/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:13

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.9
75	30.0 - 60.0% of mass 95	52.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.1
173	Less than 2.0% of mass 174	0.9 (1.3) 1
174	50.0 - 100.0% of mass 95	68.4
175	5.0 - 9.0% of mass 174	5.3 (7.8) 1
176	95.0 - 101.0% of mass 174	66.7 (97.5) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX048126.D	10/13/2025	08:55
VX1013WBL01	VX1013WBL01	VX048128.D	10/13/2025	09:46
VX1013WBS01	VX1013WBS01	VX048129.D	10/13/2025	10:43
VX1013WBSD01	VX1013WBSD01	VX048131.D	10/13/2025	11:24
P001-DS-01	Q3321-02	VX048143.D	10/13/2025	15:30