

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: LANG01

Lab Code: ACE

SDG NO.: Q3323

Continuing Calib Date: 10/13/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 10:42

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.65       | 6.66      | 6.56      | 6.76  | 0.01       |
| Aroclor-1254-2       | (2) | 6.87       | 6.88      | 6.78      | 6.98  | 0.01       |
| Aroclor-1254-3       | (3) | 7.23       | 7.24      | 7.14      | 7.34  | 0.01       |
| Aroclor-1254-4       | (4) | 7.52       | 7.52      | 7.42      | 7.62  | 0.01       |
| Aroclor-1254-5       | (5) | 7.93       | 7.94      | 7.84      | 8.04  | 0.01       |
| Tetrachloro-m-xylene |     | 4.65       | 4.66      | 4.56      | 4.76  | 0.01       |
| Decachlorobiphenyl   |     | 10.43      | 10.44     | 10.34     | 10.54 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: LANG01

Lab Code: ACE

SDG NO.: Q3323

Continuing Calib Date: 10/13/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 10:42

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.67       | 5.68      | 5.58      | 5.78 | 0.01       |
| Aroclor-1254-2 (2)   | 5.82       | 5.83      | 5.73      | 5.93 | 0.01       |
| Aroclor-1254-3 (3)   | 6.22       | 6.23      | 6.13      | 6.33 | 0.01       |
| Aroclor-1254-4 (4)   | 6.45       | 6.46      | 6.36      | 6.56 | 0.01       |
| Aroclor-1254-5 (5)   | 6.86       | 6.87      | 6.77      | 6.97 | 0.01       |
| Tetrachloro-m-xylene | 3.79       | 3.80      | 3.70      | 3.90 | 0.01       |
| Decachlorobiphenyl   | 8.79       | 8.81      | 8.71      | 8.91 | 0.02       |