

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114348.D **Time Analyzed:** 10:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.495	5.392	5.592	510.170	500.000	2.0
Aroclor-1016-2	5.516	5.414	5.614	500.630	500.000	0.1
Aroclor-1016-3	5.579	5.476	5.676	494.870	500.000	-1.0
Aroclor-1016-4	5.674	5.571	5.771	525.560	500.000	5.1
Aroclor-1016-5	5.966	5.863	6.063	516.290	500.000	3.3
Aroclor-1260-1	7.079	6.975	7.175	504.690	500.000	0.9
Aroclor-1260-2	7.333	7.229	7.429	486.340	500.000	-2.7
Aroclor-1260-3	7.689	7.585	7.785	494.550	500.000	-1.1
Aroclor-1260-4	7.911	7.807	8.007	481.320	500.000	-3.7
Aroclor-1260-5	8.218	8.113	8.313	473.220	500.000	-5.4
Decachlorobiphenyl	9.940	9.834	10.034	46.750	50.000	-6.5
Tetrachloro-m-xylene	4.357	4.255	4.455	51.380	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114348.D **Time Analyzed:** 10:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.743	4.643	4.843	536.990	500.000	7.4
Aroclor-1016-2	4.762	4.662	4.862	525.330	500.000	5.1
Aroclor-1016-3	4.937	4.836	5.036	552.520	500.000	10.5
Aroclor-1016-4	4.979	4.879	5.079	557.530	500.000	11.5
Aroclor-1016-5	5.192	5.092	5.292	530.160	500.000	6.0
Aroclor-1260-1	6.226	6.125	6.325	537.230	500.000	7.4
Aroclor-1260-2	6.414	6.313	6.513	538.050	500.000	7.6
Aroclor-1260-3	6.567	6.466	6.666	508.310	500.000	1.7
Aroclor-1260-4	7.040	6.939	7.139	542.350	500.000	8.5
Aroclor-1260-5	7.280	7.179	7.379	525.670	500.000	5.1
Decachlorobiphenyl	8.673	8.571	8.771	48.060	50.000	-3.9
Tetrachloro-m-xylene	3.660	3.560	3.760	51.500	50.000	3.0