

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114359.D **Time Analyzed:** 15:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.495	5.392	5.592	502.860	500.000	0.6
Aroclor-1016-2	5.517	5.414	5.614	503.540	500.000	0.7
Aroclor-1016-3	5.579	5.476	5.676	492.290	500.000	-1.5
Aroclor-1016-4	5.675	5.571	5.771	518.370	500.000	3.7
Aroclor-1016-5	5.966	5.863	6.063	525.650	500.000	5.1
Aroclor-1260-1	7.079	6.975	7.175	518.550	500.000	3.7
Aroclor-1260-2	7.332	7.229	7.429	489.760	500.000	-2.0
Aroclor-1260-3	7.689	7.585	7.785	494.720	500.000	-1.1
Aroclor-1260-4	7.912	7.807	8.007	479.740	500.000	-4.1
Aroclor-1260-5	8.217	8.113	8.313	471.480	500.000	-5.7
Decachlorobiphenyl	9.941	9.834	10.034	47.190	50.000	-5.6
Tetrachloro-m-xylene	4.357	4.255	4.455	52.570	50.000	5.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114359.D **Time Analyzed:** 15:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.743	4.643	4.843	539.980	500.000	8.0
Aroclor-1016-2	4.763	4.662	4.862	532.840	500.000	6.6
Aroclor-1016-3	4.938	4.836	5.036	555.940	500.000	11.2
Aroclor-1016-4	4.980	4.879	5.079	537.220	500.000	7.4
Aroclor-1016-5	5.194	5.092	5.292	513.320	500.000	2.7
Aroclor-1260-1	6.226	6.125	6.325	549.310	500.000	9.9
Aroclor-1260-2	6.414	6.313	6.513	534.560	500.000	6.9
Aroclor-1260-3	6.567	6.466	6.666	496.840	500.000	-0.6
Aroclor-1260-4	7.040	6.939	7.139	524.290	500.000	4.9
Aroclor-1260-5	7.280	7.179	7.379	499.110	500.000	-0.2
Decachlorobiphenyl	8.673	8.571	8.771	45.370	50.000	-9.3
Tetrachloro-m-xylene	3.661	3.560	3.760	51.950	50.000	3.9