

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075732.D **Time Analyzed:** 09:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.808	5.708	5.908	557.710	500.000	11.5
Aroclor-1016-2	5.830	5.729	5.929	558.720	500.000	11.7
Aroclor-1016-3	5.892	5.792	5.992	537.450	500.000	7.5
Aroclor-1016-4	5.989	5.889	6.089	580.480	500.000	16.1
Aroclor-1016-5	6.282	6.181	6.381	582.510	500.000	16.5
Aroclor-1260-1	7.400	7.299	7.499	550.730	500.000	10.1
Aroclor-1260-2	7.652	7.551	7.751	536.910	500.000	7.4
Aroclor-1260-3	8.010	7.909	8.109	579.010	500.000	15.8
Aroclor-1260-4	8.238	8.137	8.337	609.540	500.000	21.9
Aroclor-1260-5	8.564	8.464	8.664	595.760	500.000	19.2
Decachlorobiphenyl	10.435	10.334	10.534	55.030	50.000	10.1
Tetrachloro-m-xylene	4.656	4.555	4.755	59.790	50.000	19.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075732.D **Time Analyzed:** 09:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.889	4.793	4.993	567.110	500.000	13.4
Aroclor-1016-2	4.948	4.851	5.051	578.290	500.000	15.7
Aroclor-1016-3	5.067	4.971	5.171	543.170	500.000	8.6
Aroclor-1016-4	5.108	5.011	5.211	555.130	500.000	11.0
Aroclor-1016-5	5.322	5.225	5.425	542.300	500.000	8.5
Aroclor-1260-1	6.349	6.255	6.455	534.580	500.000	6.9
Aroclor-1260-2	6.537	6.441	6.641	534.700	500.000	6.9
Aroclor-1260-3	6.690	6.595	6.795	525.680	500.000	5.1
Aroclor-1260-4	7.159	7.065	7.265	472.060	500.000	-5.6
Aroclor-1260-5	7.399	7.304	7.504	493.160	500.000	-1.4
Decachlorobiphenyl	8.796	8.703	8.903	59.870	50.000	19.7
Tetrachloro-m-xylene	3.795	3.696	3.896	61.180	50.000	22.4