

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075747.D **Time Analyzed:** 16:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.805	5.708	5.908	560.430	500.000	12.1
Aroclor-1016-2	5.826	5.729	5.929	561.710	500.000	12.3
Aroclor-1016-3	5.889	5.792	5.992	544.950	500.000	9.0
Aroclor-1016-4	5.986	5.889	6.089	587.660	500.000	17.5
Aroclor-1016-5	6.278	6.181	6.381	591.140	500.000	18.2
Aroclor-1260-1	7.395	7.299	7.499	556.700	500.000	11.3
Aroclor-1260-2	7.648	7.551	7.751	549.460	500.000	9.9
Aroclor-1260-3	8.006	7.909	8.109	593.430	500.000	18.7
Aroclor-1260-4	8.233	8.137	8.337	591.710	500.000	18.3
Aroclor-1260-5	8.559	8.464	8.664	597.850	500.000	19.6
Decachlorobiphenyl	10.431	10.334	10.534	56.630	50.000	13.3
Tetrachloro-m-xylene	4.653	4.555	4.755	59.700	50.000	19.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075747.D **Time Analyzed:** 16:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.884	4.793	4.993	585.230	500.000	17.0
Aroclor-1016-2	4.944	4.851	5.051	599.380	500.000	19.9
Aroclor-1016-3	5.063	4.971	5.171	539.620	500.000	7.9
Aroclor-1016-4	5.104	5.011	5.211	577.560	500.000	15.5
Aroclor-1016-5	5.317	5.225	5.425	564.720	500.000	12.9
Aroclor-1260-1	6.346	6.255	6.455	563.870	500.000	12.8
Aroclor-1260-2	6.532	6.441	6.641	562.810	500.000	12.6
Aroclor-1260-3	6.685	6.595	6.795	548.270	500.000	9.7
Aroclor-1260-4	7.154	7.065	7.265	509.920	500.000	2.0
Aroclor-1260-5	7.395	7.304	7.504	527.660	500.000	5.5
Decachlorobiphenyl	8.791	8.703	8.903	61.140	50.000	22.3
Tetrachloro-m-xylene	3.789	3.696	3.896	59.850	50.000	19.7