

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075768.D **Time Analyzed:** 09:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.804	5.708	5.908	484.250	500.000	-3.2
Aroclor-1016-2	5.825	5.729	5.929	497.020	500.000	-0.6
Aroclor-1016-3	5.888	5.792	5.992	473.980	500.000	-5.2
Aroclor-1016-4	5.984	5.889	6.089	506.520	500.000	1.3
Aroclor-1016-5	6.277	6.181	6.381	511.230	500.000	2.2
Aroclor-1260-1	7.394	7.299	7.499	487.020	500.000	-2.6
Aroclor-1260-2	7.647	7.551	7.751	482.420	500.000	-3.5
Aroclor-1260-3	8.005	7.909	8.109	520.770	500.000	4.2
Aroclor-1260-4	8.233	8.137	8.337	536.730	500.000	7.3
Aroclor-1260-5	8.558	8.464	8.664	530.890	500.000	6.2
Decachlorobiphenyl	10.428	10.334	10.534	48.990	50.000	-2.0
Tetrachloro-m-xylene	4.651	4.555	4.755	52.680	50.000	5.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075768.D **Time Analyzed:** 09:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.886	4.793	4.993	505.240	500.000	1.0
Aroclor-1016-2	4.943	4.851	5.051	507.920	500.000	1.6
Aroclor-1016-3	5.062	4.971	5.171	465.620	500.000	-6.9
Aroclor-1016-4	5.102	5.011	5.211	493.090	500.000	-1.4
Aroclor-1016-5	5.318	5.225	5.425	462.470	500.000	-7.5
Aroclor-1260-1	6.345	6.255	6.455	480.930	500.000	-3.8
Aroclor-1260-2	6.532	6.441	6.641	471.780	500.000	-5.6
Aroclor-1260-3	6.686	6.595	6.795	463.320	500.000	-7.3
Aroclor-1260-4	7.153	7.065	7.265	435.120	500.000	-13.0
Aroclor-1260-5	7.393	7.304	7.504	456.230	500.000	-8.8
Decachlorobiphenyl	8.790	8.703	8.903	55.030	50.000	10.1
Tetrachloro-m-xylene	3.790	3.696	3.896	54.660	50.000	9.3