

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075780.D **Time Analyzed:** 13:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.803	5.708	5.908	467.090	500.000	-6.6
Aroclor-1016-2	5.825	5.729	5.929	486.420	500.000	-2.7
Aroclor-1016-3	5.888	5.792	5.992	460.820	500.000	-7.8
Aroclor-1016-4	5.985	5.889	6.089	492.570	500.000	-1.5
Aroclor-1016-5	6.277	6.181	6.381	493.370	500.000	-1.3
Aroclor-1260-1	7.394	7.299	7.499	469.060	500.000	-6.2
Aroclor-1260-2	7.647	7.551	7.751	466.250	500.000	-6.8
Aroclor-1260-3	8.005	7.909	8.109	502.500	500.000	0.5
Aroclor-1260-4	8.233	8.137	8.337	527.590	500.000	5.5
Aroclor-1260-5	8.559	8.464	8.664	511.010	500.000	2.2
Decachlorobiphenyl	10.431	10.334	10.534	47.940	50.000	-4.1
Tetrachloro-m-xylene	4.652	4.555	4.755	51.500	50.000	3.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3323
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075780.D **Time Analyzed:** 13:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.885	4.793	4.993	479.540	500.000	-4.1
Aroclor-1016-2	4.944	4.851	5.051	485.780	500.000	-2.8
Aroclor-1016-3	5.063	4.971	5.171	440.890	500.000	-11.8
Aroclor-1016-4	5.103	5.011	5.211	486.520	500.000	-2.7
Aroclor-1016-5	5.317	5.225	5.425	485.040	500.000	-3.0
Aroclor-1260-1	6.345	6.255	6.455	468.740	500.000	-6.3
Aroclor-1260-2	6.533	6.441	6.641	452.560	500.000	-9.5
Aroclor-1260-3	6.686	6.595	6.795	449.450	500.000	-10.1
Aroclor-1260-4	7.155	7.065	7.265	414.890	500.000	-17.0
Aroclor-1260-5	7.395	7.304	7.504	433.170	500.000	-13.4
Decachlorobiphenyl	8.791	8.703	8.903	54.120	50.000	8.2
Tetrachloro-m-xylene	3.791	3.696	3.896	52.210	50.000	4.4