

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3323

Client ID : SSTDCCC040

Date Analyzed: 10/14/2025

Lab File ID: BF143946.D

Time Analyzed: 13:57

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	131735	6.881	483664	8.17	238586	9.92
UPPER LIMIT	263470	7.381	967328	8.669	477172	10.422
LOWER LIMIT	65867.5	6.381	241832	7.669	119293	9.422
EPA SAMPLE NO.						
01 PB170067BL	124324	6.88	501303	8.16	273867	9.92
02 PB170067BS	118816	6.88	463801	8.17	249783	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.