

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>LANG01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3323</u>
Lab File ID:	<u>VY023424.D</u>	Date Analyzed:	<u>10/10/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>08:56</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	366006	13.346				
UPPER LIMIT	732012	13.846				
LOWER LIMIT	183003	12.846				
EPA SAMPLE NO.						
SED-01-0-2	371877	13.35				
SED-02-0-2	533404	13.35				
SED-03-0-2	408123	13.35				
SED-04-0-2	300018	13.35				
SED-05-0-2	417654	13.35				
SED-06-0-2	480884	13.35				
SED-07-0-2	372887	13.35				
VY1010SBL01	460059	13.35				
VY1010SBS01	363757	13.35				
VY1010SBSD01	356596	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.