

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3324

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038080.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	
								Qual		High	RPD
PB170061BSD	1,4-Dioxane	0.4	0.39	ug/L	98	5			70	130	20