

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG No.: Q3324

Instrument ID: MSVOA_X

Calibration Date/Time: 10/10/2025 10:45

Lab File ID: VX048100.D

Init. Calib. Date(s): 09/16/2025 09/16/2025

Heated Purge: (Y/N) N

Init. Calib. Time(s): 09:23 12:01

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.649		25.29	20
Chloromethane	0.680	0.784	0.1	15.29	20
Vinyl Chloride	0.666	0.823		23.57	20
Bromomethane	0.422	0.549		30.09	20
Chloroethane	0.445	0.521		17.08	20
Trichlorofluoromethane	0.989	1.220		23.36	20
1,1,2-Trichlorotrifluoroethane	0.578	0.696		20.42	20
1,1-Dichloroethene	0.594	0.700		17.84	20
Acetone	0.346	0.329		-4.91	20
Carbon Disulfide	1.626	2.098		29.03	20
Methyl tert-butyl Ether	2.169	2.245		3.5	20
Methyl Acetate	0.770	0.680		-11.69	20
Methylene Chloride	0.732	0.797		8.88	20
trans-1,2-Dichloroethene	0.640	0.750		17.19	20
1,1-Dichloroethane	1.263	1.418	0.1	12.27	20
Cyclohexane	1.052	1.209		14.92	20
2-Butanone	0.432	0.394		-8.8	20
Carbon Tetrachloride	0.532	0.618		16.17	20
cis-1,2-Dichloroethene	0.783	0.864		10.35	20
Bromochloromethane	0.551	0.545		-1.09	20
Chloroform	1.276	1.444		13.17	20
1,1,1-Trichloroethane	1.082	1.249		15.43	20
Methylcyclohexane	0.556	0.683		22.84	20
Benzene	1.488	1.737		16.73	20
1,2-Dichloroethane	0.566	0.633		11.84	20
Trichloroethene	0.360	0.428		18.89	20
1,2-Dichloropropane	0.381	0.437		14.7	20
Bromodichloromethane	0.572	0.653		14.16	20
4-Methyl-2-Pentanone	0.477	0.460		-3.56	20
Toluene	0.917	1.068		16.47	20
t-1,3-Dichloropropene	0.584	0.658		12.67	20
cis-1,3-Dichloropropene	0.624	0.727		16.51	20
1,1,2-Trichloroethane	0.354	0.390		10.17	20
2-Hexanone	0.343	0.326		-4.96	20
Dibromochloromethane	0.407	0.478		17.44	20
1,2-Dibromoethane	0.360	0.413		14.72	20
Tetrachloroethene	0.326	0.385		18.1	20
Chlorobenzene	1.136	1.300	0.3	14.44	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3324</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/10/2025 10:45</u>
Lab File ID:	<u>VX048100.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.276		16.12	20
m/p-Xylenes	0.729	0.856		17.42	20
o-Xylene	0.706	0.816		15.58	20
Styrene	1.236	1.415		14.48	20
Bromoform	0.293	0.318	0.1	8.53	20
Isopropylbenzene	3.801	4.257		12	20
1,1,2,2-Tetrachloroethane	1.150	1.154	0.3	0.35	20
1,3-Dichlorobenzene	1.739	1.911		9.89	20
1,4-Dichlorobenzene	1.785	1.930		8.12	20
1,2-Dichlorobenzene	1.657	1.793		8.21	20
1,2-Dibromo-3-Chloropropane	0.233	0.211		-9.44	20
1,2,4-Trichlorobenzene	1.077	1.124		4.36	20
1,2,3-Trichlorobenzene	1.002	1.066		6.39	20
1,2-Dichloroethane-d4	0.796	0.748		-6.03	20
Dibromofluoromethane	0.335	0.343		2.39	20
Toluene-d8	1.160	1.162		0.17	20
4-Bromofluorobenzene	0.440	0.449		2.05	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.