

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG No.: Q3324

Instrument ID: MSVOA_X

Calibration Date/Time: 10/10/2025 19:40

Lab File ID: VX048124.D

Init. Calib. Date(s): 09/16/2025 09/16/2025

Heated Purge: (Y/N) N

Init. Calib. Time(s): 09:23 12:01

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.617		19.11	50
Chloromethane	0.680	0.743	0.1	9.27	50
Vinyl Chloride	0.666	0.797		19.67	50
Bromomethane	0.422	0.510		20.85	50
Chloroethane	0.445	0.508		14.16	50
Trichlorofluoromethane	0.989	1.152		16.48	50
1,1,2-Trichlorotrifluoroethane	0.578	0.650		12.46	50
1,1-Dichloroethene	0.594	0.679		14.31	50
Acetone	0.346	0.280		-19.08	50
Carbon Disulfide	1.626	1.966		20.91	50
Methyl tert-butyl Ether	2.169	2.347		8.21	50
Methyl Acetate	0.770	0.721		-6.36	50
Methylene Chloride	0.732	0.826		12.84	50
trans-1,2-Dichloroethene	0.640	0.741		15.78	50
1,1-Dichloroethane	1.263	1.427	0.1	12.98	50
Cyclohexane	1.052	1.175		11.69	50
2-Butanone	0.432	0.424		-1.85	50
Carbon Tetrachloride	0.532	0.566		6.39	50
cis-1,2-Dichloroethene	0.783	0.891		13.79	50
Bromochloromethane	0.551	0.649		17.79	50
Chloroform	1.276	1.469		15.13	50
1,1,1-Trichloroethane	1.082	1.227		13.4	50
Methylcyclohexane	0.556	0.594		6.84	50
Benzene	1.488	1.653		11.09	50
1,2-Dichloroethane	0.566	0.617		9.01	50
Trichloroethene	0.360	0.392		8.89	50
1,2-Dichloropropane	0.381	0.421		10.5	50
Bromodichloromethane	0.572	0.631		10.31	50
4-Methyl-2-Pentanone	0.477	0.494		3.56	50
Toluene	0.917	1.029		12.21	50
t-1,3-Dichloropropene	0.584	0.624		6.85	50
cis-1,3-Dichloropropene	0.624	0.688		10.26	50
1,1,2-Trichloroethane	0.354	0.387		9.32	50
2-Hexanone	0.343	0.337		-1.75	50
Dibromochloromethane	0.407	0.467		14.74	50
1,2-Dibromoethane	0.360	0.412		14.44	50
Tetrachloroethene	0.326	0.347		6.44	50
Chlorobenzene	1.136	1.198	0.3	5.46	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3324</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/10/2025 19:40</u>
Lab File ID:	<u>VX048124.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.041		4.13	50
m/p-Xylenes	0.729	0.782		7.27	50
o-Xylene	0.706	0.740		4.82	50
Styrene	1.236	1.307		5.74	50
Bromoform	0.293	0.299	0.1	2.05	50
Isopropylbenzene	3.801	3.460		-8.97	50
1,1,2,2-Tetrachloroethane	1.150	1.042	0.3	-9.39	50
1,3-Dichlorobenzene	1.739	1.598		-8.11	50
1,4-Dichlorobenzene	1.785	1.614		-9.58	50
1,2-Dichlorobenzene	1.657	1.535		-7.36	50
1,2-Dibromo-3-Chloropropane	0.233	0.200		-14.16	50
1,2,4-Trichlorobenzene	1.077	0.936		-13.09	50
1,2,3-Trichlorobenzene	1.002	0.875		-12.68	50
1,2-Dichloroethane-d4	0.796	0.857		7.66	50
Dibromofluoromethane	0.335	0.345		2.98	50
Toluene-d8	1.160	1.159		-0.09	50
4-Bromofluorobenzene	0.440	0.478		8.64	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.