



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101325\
Data File : VX048125.D
Acq On : 13 Oct 2025 08:13
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Title : SW846 8260
Last Update : Wed Sep 17 06:39:58 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1628

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	11954	PASS
75	95	30	60	52.4	31523	PASS
95	95	100	100	100.0	60112	PASS
96	95	5	9	6.1	3687	PASS
173	174	0.00	2	1.3	529	PASS
174	95	50	100	68.4	41144	PASS
175	174	5	9	7.8	3196	PASS
176	174	95	101	97.5	40099	PASS
177	176	5	9	6.9	2756	PASS