

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX091625\  
 Data File : VX047589.D  
 Acq On : 16 Sep 2025 11:40  
 Operator : JC/MD  
 Sample : VSTDICC100  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDICC100

Quant Time: Sep 17 06:15:00 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 17 06:04:35 2025  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By : John Carlone 09/17/2025  
 Supervised By : Mahesh Dadoda 09/18/2025

| Compound                     | R.T.           | QIon | Response             | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|----------------------|---------|--------|----------|
| -----                        |                |      |                      |         |        |          |
| Internal Standards           |                |      |                      |         |        |          |
| 1) Pentafluorobenzene        | 5.532          | 168  | 169280               | 50.000  | ug/l   | -0.01    |
| 34) 1,4-Difluorobenzene      | 6.739          | 114  | 298913               | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.037         | 117  | 258399               | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.006         | 152  | 122218               | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |                      |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.934          | 65   | 275758               | 102.341 | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = 204.680%# |         |        |          |
| 35) Dibromofluoromethane     | 5.361          | 113  | 212419               | 105.962 | ug/l   | -0.01    |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = 211.920%# |         |        |          |
| 50) Toluene-d8               | 8.629          | 98   | 724078               | 104.386 | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = 208.780%# |         |        |          |
| 62) 4-Bromofluorobenzene     | 11.061         | 95   | 270289               | 102.672 | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = 205.340%# |         |        |          |
| Target Compounds             |                |      |                      |         |        |          |
|                              |                |      |                      |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.173          | 85   | 172727               | 98.537  | ug/l   | 100      |
| 3) Chloromethane             | 1.301          | 50   | 229165               | 99.472  | ug/l   | 100      |
| 4) Vinyl Chloride            | 1.380          | 62   | 229175               | 101.621 | ug/l   | 100      |
| 5) Bromomethane              | 1.611          | 94   | 143113               | 100.076 | ug/l   | 99       |
| 6) Chloroethane              | 1.691          | 64   | 152487               | 101.167 | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 1.892          | 101  | 344488               | 102.878 | ug/l   | 100      |
| 8) Diethyl Ether             | 2.130          | 74   | 142038               | 103.695 | ug/l   | 100      |
| 9) 1,1,2-Trichlorotrifluo... | 2.331          | 101  | 199843               | 102.067 | ug/l   | 99       |
| 10) Methyl Iodide            | 2.453          | 142  | 313092               | 102.531 | ug/l   | 99       |
| 11) Tert butyl alcohol       | 2.947          | 59   | 152108               | 506.863 | ug/l   | 98       |
| 12) 1,1-Dichloroethene       | 2.319          | 96   | 206097               | 102.473 | ug/l   | 93       |
| 13) Acrolein                 | 2.233          | 56   | 150793               | 545.276 | ug/l   | 97       |
| 14) Allyl chloride           | 2.660          | 41   | 415140               | 100.028 | ug/l   | 99       |
| 15) Acrylonitrile            | 3.056          | 53   | 581240               | 522.897 | ug/l   | 100      |
| 16) Acetone                  | 2.374          | 43   | 510542               | 435.772 | ug/l   | 100      |
| 17) Carbon Disulfide         | 2.514          | 76   | 537888               | 97.693  | ug/l   | 98       |
| 18) Methyl Acetate           | 2.697          | 43   | 292867               | 112.324 | ug/l   | 100      |
| 19) Methyl tert-butyl Ether  | 3.105          | 73   | 748952               | 102.011 | ug/l   | 100      |
| 20) Methylene Chloride       | 2.782          | 84   | 242548               | 97.805  | ug/l   | 97       |
| 21) trans-1,2-Dichloroethene | 3.087          | 96   | 217759               | 100.566 | ug/l   | 98       |
| 22) Diisopropyl ether        | 3.745          | 45   | 811840               | 100.934 | ug/l   | 88       |
| 23) Vinyl Acetate            | 3.709          | 43   | 3314707              | 514.916 | ug/l   | 99       |
| 24) 1,1-Dichloroethane       | 3.599          | 63   | 435815               | 101.944 | ug/l   | 98       |
| 25) 2-Butanone               | 4.532          | 43   | 700827               | 479.558 | ug/l   | 97       |
| 26) 2,2-Dichloropropane      | 4.465          | 77   | 357292               | 100.542 | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 4.471          | 96   | 266550               | 100.516 | ug/l   | 100      |
| 28) Bromochloromethane       | 4.873          | 49   | 198788               | 106.512 | ug/l   | 99       |
| 29) Tetrahydrofuran          | 4.977          | 42   | 443530               | 504.502 | ug/l   | 99       |
| 30) Chloroform               | 5.074          | 83   | 433937               | 100.447 | ug/l   | 100      |
| 31) Cyclohexane              | 5.452          | 56   | 332995               | 93.523  | ug/l   | 96       |
| 32) 1,1,1-Trichloroethane    | 5.367          | 97   | 372323               | 101.607 | ug/l   | 99       |
| 36) 1,1-Dichloropropene      | 5.672          | 75   | 281650               | 96.149  | ug/l   | 98       |
| 37) Ethyl Acetate            | 4.690          | 43   | 318793               | 101.641 | ug/l   | 99       |
| 38) Carbon Tetrachloride     | 5.660          | 117  | 314745               | 98.891  | ug/l   | 98       |
| 39) Methylcyclohexane        | 7.360          | 83   | 327858               | 98.600  | ug/l   | 98       |
| 40) Benzene                  | 6.019          | 78   | 880885               | 99.009  | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX091625\  
 Data File : VX047589.D  
 Acq On : 16 Sep 2025 11:40  
 Operator : JC/MD  
 Sample : VSTDICC100  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDICC100

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 09/17/2025  
 Supervised By : Mahesh Dadoda 09/18/2025

Quant Time: Sep 17 06:15:00 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 17 06:04:35 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|------|----------|----------|-------|----------|
| 41) Methacrylonitrile         | 4.891  | 41   | 168753   | 104.220  | ug/l  | 99       |
| 42) 1,2-Dichloroethane        | 6.062  | 62   | 331465   | 97.916   | ug/l  | 100      |
| 43) Isopropyl Acetate         | 6.318  | 43   | 516875   | 100.482  | ug/l  | 99       |
| 44) Trichloroethene           | 7.104  | 130  | 217753   | 101.234  | ug/l  | 96       |
| 45) 1,2-Dichloropropane       | 7.409  | 63   | 228150   | 100.145  | ug/l  | 100      |
| 46) Dibromomethane            | 7.562  | 93   | 164987   | 99.519   | ug/l  | 99       |
| 47) Bromodichloromethane      | 7.806  | 83   | 345991   | 101.136  | ug/l  | 98       |
| 48) Methyl methacrylate       | 7.671  | 41   | 260358   | 101.612  | ug/l  | 100      |
| 49) 1,4-Dioxane               | 7.641  | 88   | 56717    | 2190.315 | ug/l  | 99       |
| 51) 4-Methyl-2-Pentanone      | 8.555  | 43   | 1431253  | 501.834  | ug/l  | 100      |
| 52) Toluene                   | 8.702  | 92   | 536148   | 97.811   | ug/l  | 99       |
| 53) t-1,3-Dichloropropene     | 8.958  | 75   | 355881   | 101.986  | ug/l  | 100      |
| 54) cis-1,3-Dichloropropene   | 8.348  | 75   | 380577   | 102.038  | ug/l  | 99       |
| 55) 1,1,2-Trichloroethane     | 9.135  | 97   | 205939   | 97.434   | ug/l  | 97       |
| 56) Ethyl methacrylate        | 9.098  | 69   | 346659   | 104.435  | ug/l  | 99       |
| 57) 1,3-Dichloropropane       | 9.293  | 76   | 362848   | 99.935   | ug/l  | 98       |
| 58) 2-Chloroethyl Vinyl ether | 8.226  | 63   | 836362   | 499.814  | ug/l  | 100      |
| 59) 2-Hexanone                | 9.415  | 43   | 996274   | 486.380  | ug/l  | 100      |
| 60) Dibromochloromethane      | 9.506  | 129  | 251394   | 103.226  | ug/l  | 99       |
| 61) 1,2-Dibromoethane         | 9.592  | 107  | 217751   | 101.147  | ug/l  | 100      |
| 64) Tetrachloroethene         | 9.256  | 164  | 165559   | 98.284   | ug/l  | 98       |
| 65) Chlorobenzene             | 10.061 | 112  | 588517   | 100.253  | ug/l  | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 10.147 | 131  | 208621   | 102.988  | ug/l  | 100      |
| 67) Ethyl Benzene             | 10.177 | 91   | 1031346  | 101.823  | ug/l  | 100      |
| 68) m/p-Xylenes               | 10.287 | 106  | 763898   | 202.624  | ug/l  | 99       |
| 69) o-Xylene                  | 10.622 | 106  | 372621   | 102.186  | ug/l  | 99       |
| 70) Styrene                   | 10.634 | 104  | 649863   | 101.710  | ug/l  | 100      |
| 71) Bromoform                 | 10.781 | 173  | 161188   | 106.566  | ug/l  | 98       |
| 73) Isopropylbenzene          | 10.945 | 105  | 954900   | 102.768  | ug/l  | 100      |
| 74) N-amyl acetate            | 10.823 | 43   | 448344   | 105.812  | ug/l  | 99       |
| 75) 1,1,2,2-Tetrachloroethane | 11.195 | 83   | 281608   | 100.172  | ug/l  | 98       |
| 76) 1,2,3-Trichloropropane    | 11.220 | 75   | 229521m  | 90.444   | ug/l  |          |
| 77) Bromobenzene              | 11.183 | 156  | 234666   | 102.541  | ug/l  | 98       |
| 78) n-propylbenzene           | 11.287 | 91   | 1119728  | 101.941  | ug/l  | 99       |
| 79) 2-Chlorotoluene           | 11.348 | 91   | 683867   | 101.652  | ug/l  | 99       |
| 80) 1,3,5-Trimethylbenzene    | 11.433 | 105  | 787327   | 103.134  | ug/l  | 100      |
| 81) trans-1,4-Dichloro-2-b... | 11.000 | 75   | 102555   | 105.689  | ug/l  | 98       |
| 82) 4-Chlorotoluene           | 11.433 | 91   | 802929   | 101.054  | ug/l  | 100      |
| 83) tert-Butylbenzene         | 11.695 | 119  | 791896   | 101.333  | ug/l  | 100      |
| 84) 1,2,4-Trimethylbenzene    | 11.732 | 105  | 790557   | 102.144  | ug/l  | 99       |
| 85) sec-Butylbenzene          | 11.872 | 105  | 945321   | 101.840  | ug/l  | 100      |
| 86) p-Isopropyltoluene        | 11.994 | 119  | 797643   | 101.501  | ug/l  | 99       |
| 87) 1,3-Dichlorobenzene       | 11.951 | 146  | 432248   | 101.689  | ug/l  | 100      |
| 88) 1,4-Dichlorobenzene       | 12.024 | 146  | 431839   | 98.985   | ug/l  | 100      |
| 89) n-Butylbenzene            | 12.311 | 91   | 757239   | 104.138  | ug/l  | 99       |
| 90) Hexachloroethane          | 12.518 | 117  | 148048   | 107.304  | ug/l  | 100      |
| 91) 1,2-Dichlorobenzene       | 12.317 | 146  | 414491   | 102.352  | ug/l  | 100      |
| 92) 1,2-Dibromo-3-Chloropr... | 12.920 | 75   | 61029    | 107.225  | ug/l  | 98       |
| 93) 1,2,4-Trichlorobenzene    | 13.567 | 180  | 278777   | 105.922  | ug/l  | 100      |
| 94) Hexachlorobutadiene       | 13.707 | 225  | 92124    | 103.049  | ug/l  | 99       |
| 95) Naphthalene               | 13.756 | 128  | 854492   | 106.620  | ug/l  | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.939 | 180  | 263512   | 107.633  | ug/l  | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX091625\  
Data File : VX047589.D  
Acq On : 16 Sep 2025 11:40  
Operator : JC/MD  
Sample : VSTDICC100  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDICC100

Manual Integrations  
APPROVED

Reviewed By :John Carlone 09/17/2025  
Supervised By :Mahesh Dadoda 09/18/2025

Quant Time: Sep 17 06:15:00 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 17 06:04:35 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX091625\  
Data File : VX047589.D  
Acq On : 16 Sep 2025 11:40  
Operator : JC/MD  
Sample : VSTDIC100  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDIC100

Manual Integrations  
APPROVED

Reviewed By : John Carlone 09/17/2025  
Supervised By : Mahesh Dadoda 09/18/2025

Quant Time: Sep 17 06:15:00 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 17 06:04:35 2025  
Response via : Initial Calibration

