

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG NO.: Q3325

Lab File ID: VX047584.D

BFB Injection Date: 09/16/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:56

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20
75	30.0 - 60.0% of mass 95	54
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	67.3
175	5.0 - 9.0% of mass 174	5.4 (8.1) 1
176	95.0 - 101.0% of mass 174	66.3 (98.6) 1
177	5.0 - 9.0% of mass 176	4.3 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX047585.D	09/16/2025	09:23
VSTDIC005	VSTDIC005	VX047586.D	09/16/2025	10:16
VSTDIC020	VSTDIC020	VX047587.D	09/16/2025	10:37
VSTDIC050	VSTDIC050	VX047588.D	09/16/2025	10:58
VSTDIC100	VSTDIC100	VX047589.D	09/16/2025	11:40
VSTDIC150	VSTDIC150	VX047590.D	09/16/2025	12:01