

### Initial and Continuing Calibration Verification

**Client:** Europastry

**SDG No.:** Q3328

**Project:** MCUA Permit No 14241 - 571 Jersey Ave NB NJ

**RunNo.:** LB137541

| Analyte           |      | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|-------------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID:<br>COD | ICV  | mg/L  | 50.173 | 50         | 100           | 95-105                    | 09/08/2025       |
| Sample ID:<br>COD | CCV1 | mg/L  | 48.150 | 50         | 96            | 95-105                    | 10/15/2025       |
| Sample ID:<br>COD | CCV2 | mg/L  | 51.184 | 50         | 102           | 95-105                    | 10/15/2025       |

## Initial and Continuing Calibration Verification

**Client:** Europastry

**SDG No.:** Q3328

**Project:** MCUA Permit No 14241 - 571 Jersey Ave NB NJ

**RunNo.:** LB137551

| Analyte                      | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV</b><br>pH  | pH    | 6.98   | 7          | 100        | 90-110                 | 10/16/2025    |
| Sample ID: <b>CCV1</b><br>pH | pH    | 2.01   | 2.00       | 101        | 90-110                 | 10/16/2025    |
| Sample ID: <b>CCV2</b><br>pH | pH    | 12.02  | 12.00      | 100        | 90-110                 | 10/16/2025    |