

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3328
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	LB137557BS
Client ID:	LB137557BSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	17.0		17.1		1	0.59		10/16/2025

Duplicate Sample Summary

Client: Europastry Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ Client ID: COMPDUP	SDG No.: Q3328 Sample ID: Q3316-02 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	300		301		1	0.33		10/14/2025

Duplicate Sample Summary

Client: Europastry	SDG No.: Q3328
Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID: Q3328-01
Client ID: MH-10-10-25DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	472		473		1	0.25		10/10/2025
pH	pH	+/-20	4.88		4.90		1	0.41		10/16/2025

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3328
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3328-01
Client ID:	MH-10-10-25MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	196		196		1	0.05		10/15/2025

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3328
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3345-07
Client ID:	TANK-FARM-WATERDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	5.67	J	5.67	J	1	0		10/15/2025

Duplicate Sample Summary

Client: Europastry Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ Client ID: TANK-FARM-WATERMSD	SDG No.: Q3328 Sample ID: Q3345-07 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	52.2		53.2		1	1.9		10/15/2025

Duplicate Sample Summary

Client: Europastry	SDG No.: Q3328
Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID: Q3356-01
Client ID: AUD-25-0169-AUD-25-0170DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	7.20		7.22		1	0.28		10/16/2025