

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3328
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137496

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137496BS							
BOD5	mg/L	198	184		93	1	84.6-115.4	10/10/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3328
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137515

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137515BS							
TSS	mg/L	550	533		97	1	90-110	10/14/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3328
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137541

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137541BS							
COD	mg/L	50	51.2		102	1	90-110	10/15/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3328
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137556

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137556BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	10/15/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3328
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137557

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137557BS							
TPH	mg/L	20.0	17.0		85	1	78-114	10/16/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3328
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137557

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137557BSD							
TPH	mg/L	20.0	17.1		86	1	78-114	10/16/2025