

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc.

SDG No.: Q3335

Contract: TOTA02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	4090	4000	102	90 - 110	P	10/17/2025	12:33	LB137588

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc.

SDG No.: Q3335

Contract: TOTA02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4820	5000	96	90 - 110	P	10/17/2025	13:23	LB137588
CCV02	Lead	5100	5000	102	90 - 110	P	10/17/2025	14:21	LB137588

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc.

SDG No.: Q3335

Contract: TOTA02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	13.2	12.0	110	80 - 120	P	10/17/2025	14:26	LB137588

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Total Quality Environmental Inc.

SDG No.: Q3335

Contract: TOTA02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV03	Lead	4950	5000	99	90 - 110	P	10/17/2025	15:17	LB137588
CCV04	Lead	5080	5000	102	90 - 110	P	10/17/2025	16:07	LB137588
CCV05	Lead	5070	5000	102	90 - 110	P	10/17/2025	16:59	LB137588
CCV06	Lead	4940	5000	99	90 - 110	P	10/17/2025	17:41	LB137588