



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Metals

- 3a -

#### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

**Client:** Total Quality Environmental Inc.

**SDG No.:** Q3335

**Contract:** TOTA02

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| ICB01     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 10/17/2025       | 12:49            | LB137588      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** Total Quality Environmental Inc.

**SDG No.:** Q3335

**Contract:** TOTA02

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| CCB01     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 10/17/2025       | 13:28            | LB137588      |
| CCB02     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 10/17/2025       | 14:30            | LB137588      |
| CCB03     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 10/17/2025       | 15:21            | LB137588      |
| CCB04     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 10/17/2025       | 16:11            | LB137588      |
| CCB05     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 10/17/2025       | 17:03            | LB137588      |
| CCB06     | Lead    | 2.30           | +/-6                | U            | 12.0 | P | 10/17/2025       | 17:45            | LB137588      |