

**Metals**

- 4 -

**INTERFERENCE CHECK SAMPLE**

**Client:** Total Quality Environmental Inc.

**SDG No.:** Q3335

**Contract:** TOTA02

**Lab Code:** ACE

**ICS Source:** EPA

**Instrument ID:** P4

| Sample ID | Analyte | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Low<br>Limit<br>(ug/L) | High<br>Limit<br>(ug/L) | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|--------------------|---------------|------------------------|-------------------------|------------------|------------------|---------------|
| ICSA01    | Lead    | 2.85           |                    |               | -12                    | 12                      | 10/17/2025       | 13:01            | LB137588      |
| ICSAB01   | Lead    | 52.4           | 49.0               | 107           | 37                     | 61                      | 10/17/2025       | 13:06            | LB137588      |