

metals  
- 5a -  
**MATRIX SPIKE SUMMARY**

**client:** Total Quality Environmental Inc.      **level:** low      **sdg no.:** Q3335  
**contract:** TOTA02      **lab code:** ACE  
**matrix:** Water      **sample id:** Q3335-14      **client id:** R (2nd draw) MS  
**Percent Solids for Sample:** NA      **Spiked ID:** Q3335-14MS      **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	460		35.1		500	85		P

metals  
- 5a -  
**MATRIX SPIKE DUPLICATE SUMMARY**

**client:** Total Quality Environmental Inc.      **level:** low      **sdg no.:** Q3335  
**contract:** TOTA02      **lab code:** ACE  
**matrix:** Water      **sample id:** Q3335-14      **client id:** R (2nd draw) MSD  
**Percent Solids for Sample:** NA      **Spiked ID:** Q3335-14MSD      **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	491		35.1		500	91		P