

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

**Client:** Total Quality Environmental Inc.

**SDG No.:** Q3335

**Contract:** TOTA02

**Lab Code:** ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170086BS Lead	ug/L	500	444		89	80 - 120	P