

Initial and Continuing Calibration Verification

Client: Scientific Design
Project: Carrier Post-Treatment

SDG No.: Q3339
RunNo.: LB137529

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TOC	mg/L	10.2	10	102	90-110	10/06/2025
Sample ID: CCV1 TOC	mg/L	9.8	10	98	90-110	10/15/2025
Sample ID: CCV2 TOC	mg/L	10	10	100	90-110	10/15/2025
Sample ID: CCV3 TOC	mg/L	10.2	10	102	90-110	10/15/2025