

Laboratory Control Sample Summary

Client: Kleinfelder
Project: Webster School

SDG No.: Q3341
Run No.: LB137519

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137519BSS							
Bromide	mg/Kg	200	204		102	1	90-110	10/15/2025
Chloride	mg/Kg	60	61.0		102	1	90-110	10/15/2025
Fluoride	mg/Kg	40	40.8		102	1	90-110	10/15/2025
Nitrite	mg/Kg	60	60.6		101	1	90-110	10/15/2025
Nitrate	mg/Kg	50	49.6		99	1	90-110	10/15/2025
Sulfate	mg/Kg	300	303		101	1	90-110	10/15/2025
Orthophosphate as P	mg/Kg	100	104		104	1	90-110	10/15/2025

Laboratory Control Sample Summary

Client: Kleinfelder
Project: Webster School

SDG No.: Q3341
Run No.: LB137525

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170093BS							
Hexavalent Chromium	mg/Kg	20	19.7		98	1	84-110	10/15/2025

Laboratory Control Sample Summary

Client: Kleinfelder
Project: Webster School

SDG No.: Q3341
Run No.: LB137539

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170110BS							
Ammonia as N	mg/Kg	50	50.0		100	1	90-110	10/15/2025