

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137528

Review By	Sagar	Review On	10/16/2025 12:12:58 PM
Supervise By	jaswal	Supervise On	10/16/2025 11:28:01 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87584,MP87585,MP87586,MP87587,MP87588,MP87589,MP87590		
ICV Standard	MP87591		
CCV Standard	MP87593		
ICSA Standard			
CRI Standard	MP87595		
LCS Standard			
Chk Standard	MP87592,MP87594,MP87596,MP87598		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/15/25 08:47		Ayul	OK
2	S0.2	S0.2	CAL2	10/15/25 08:49		Ayul	OK
3	S2.5	S2.5	CAL3	10/15/25 08:51		Ayul	OK
4	S5	S5	CAL4	10/15/25 08:54		Ayul	OK
5	S7.5	S7.5	CAL5	10/15/25 08:56		Ayul	OK
6	S10	S10	CAL6	10/15/25 08:59		Ayul	OK
7	ICV71	ICV71	ICV	10/15/25 09:02		Ayul	OK
8	ICB71	ICB71	ICB	10/15/25 09:08		Ayul	OK
9	CCV29	CCV29	CCV	10/15/25 09:10		Ayul	OK
10	CCB29	CCB29	CCB	10/15/25 09:12		Ayul	OK
11	HighStd	HighStd	HIGH STD	10/15/25 09:17		Ayul	OK
12	ChkStd	ChkStd	SAM	10/15/25 09:20		Ayul	OK
13	PB170097BL	PB170097BL	MB	10/15/25 09:22		Ayul	OK
14	PB170097BS	PB170097BS	LCS	10/15/25 09:28		Ayul	OK
15	Q3337-01	LAYDOWN-YARD-1	SAM	10/15/25 09:30		Ayul	OK
16	Q3337-03	LAYDOWN-YARD-2	SAM	10/15/25 09:33		Ayul	OK
17	Q3341-01	C-1	SAM	10/15/25 09:35		Ayul	OK
18	Q3341-02	C-2	SAM	10/15/25 09:37		Ayul	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137528

Review By	Sagar	Review On	10/16/2025 12:12:58 PM
Supervise By	jaswal	Supervise On	10/16/2025 11:28:01 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87584,MP87585,MP87586,MP87587,MP87588,MP87589,MP87590		
ICV Standard	MP87591		
CCV Standard	MP87593		
ICSA Standard			
CRI Standard	MP87595		
LCS Standard			
Chk Standard	MP87592,MP87594,MP87596,MP87598		

19	Q3341-03	C-3	SAM	10/15/25 09:40		Ayul	OK
20	CCV30	CCV30	CCV	10/15/25 09:42		Ayul	OK
21	CCB30	CCB30	CCB	10/15/25 09:44		Ayul	OK
22	Q3342-01	CONC-ROLLOFF-WC	SAM	10/15/25 09:47		Ayul	OK
23	Q3343-01	SOIL-ROLLOFF-WC-	SAM	10/15/25 09:49		Ayul	OK
24	Q3343-01DUP	SOIL-ROLLOFF-WC-	DUP	10/15/25 09:52		Ayul	OK
25	Q3343-01MS	SOIL-ROLLOFF-WC-	MS	10/15/25 09:54		Ayul	OK
26	Q3343-01MSD	SOIL-ROLLOFF-WC-	MSD	10/15/25 10:03		Ayul	OK
27	Q3345-01	TANK-FARM-SOIL	SAM	10/15/25 10:05		Ayul	OK
28	Q3345-03	2ND-STREET-SOIL	SAM	10/15/25 10:08		Ayul	OK
29	Q3345-05	M-R-SOIL	SAM	10/15/25 10:10		Ayul	OK
30	Q3343-01L	SOIL-ROLLOFF-WC-	SD	10/15/25 10:13		Ayul	OK
31	Q3343-01A	SOIL-ROLLOFF-WC-	PS	10/15/25 10:15		Ayul	OK
32	CCV31	CCV31	CCV	10/15/25 10:18		Ayul	OK
33	CCB31	CCB31	CCB	10/15/25 10:23		Ayul	OK
34	CRA	CRA	CRDL	10/15/25 10:27		Ayul	OK
35	CCV32	CCV32	CCV	10/15/25 10:30		Ayul	OK
36	CCB32	CCB32	CCB	10/15/25 10:32		Ayul	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137561

Review By	Janvi	Review On	10/17/2025 12:20:45 PM
Supervise By	jaswal	Supervise On	10/17/2025 1:18:42 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87404,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87405		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/15/25 13:15		Jaswal	OK
2	S1	S1	CAL2	10/15/25 13:19		Jaswal	OK
3	S2	S2	CAL3	10/15/25 13:24		Jaswal	OK
4	S3	S3	CAL4	10/15/25 13:28		Jaswal	OK
5	S4	S4	CAL5	10/15/25 13:32		Jaswal	OK
6	S5	S5	CAL6	10/15/25 13:36		Jaswal	OK
7	ICV01	ICV01	ICV	10/15/25 14:07		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/15/25 14:19		Jaswal	OK
9	ICB01	ICB01	ICB	10/15/25 14:24		Jaswal	OK
10	CRI01	CRI01	CRDL	10/15/25 14:28		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/15/25 14:39		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/15/25 14:52		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/15/25 14:56		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/15/25 15:00		Jaswal	OK
15	CCV01	CCV01	CCV	10/15/25 15:09		Jaswal	OK
16	CCB01	CCB01	CCB	10/15/25 15:13		Jaswal	OK
17	LR-2	LR-2	HIGH STD	10/15/25 15:22		Jaswal	OK
18	LR-3	LR-3	HIGH STD	10/15/25 16:06		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137561

Review By	Janvi	Review On	10/17/2025 12:20:45 PM
Supervise By	jaswal	Supervise On	10/17/2025 1:18:42 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87404,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87405		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

19	LR-1	LR-1	HIGH STD	10/15/25 16:12		Jaswal	OK
20	CCV02	CCV02	CCV	10/15/25 16:28		Jaswal	OK
21	CCB02	CCB02	CCB	10/15/25 16:32		Jaswal	OK
22	PB170089BL	PB170089BL	MB	10/15/25 16:36		Jaswal	OK
23	PB170089BS	PB170089BS	LCS	10/15/25 16:43		Jaswal	OK
24	Q3337-01	LAYDOWN-YARD-1	SAM	10/15/25 16:47		Jaswal	OK
25	Q3337-03	LAYDOWN-YARD-2	SAM	10/15/25 16:51		Jaswal	OK
26	Q3337-03DUP	LAYDOWN-YARD-2D	DUP	10/15/25 16:55		Jaswal	OK
27	Q3337-03L	LAYDOWN-YARD-2L	SD	10/15/25 16:59		Jaswal	OK
28	Q3337-03MS	LAYDOWN-YARD-2M	MS	10/15/25 17:04		Jaswal	OK
29	Q3337-03MSD	LAYDOWN-YARD-2M	MSD	10/15/25 17:08		Jaswal	OK
30	Q3337-03A	LAYDOWN-YARD-2A	PS	10/15/25 17:12	0.1 ml of each M6008 AND M6017 IN 10ML SAMPLE.	Jaswal	OK
31	PB170108BL	PB170108BL	MB	10/15/25 17:16		Jaswal	OK
32	CCV03	CCV03	CCV	10/15/25 17:21		Jaswal	OK
33	CCB03	CCB03	CCB	10/15/25 17:25		Jaswal	OK
34	PB170108BS	PB170108BS	LCS	10/15/25 17:29		Jaswal	OK
35	Q3345-07	TANK-FARM-WATER	SAM	10/15/25 17:33		Jaswal	OK
36	Q3345-07DUP	TANK-FARM-WATER	DUP	10/15/25 17:37		Jaswal	OK
37	Q3345-07L	TANK-FARM-WATER	SD	10/15/25 17:41		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137561

Review By	Janvi	Review On	10/17/2025 12:20:45 PM
Supervise By	jaswal	Supervise On	10/17/2025 1:18:42 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87404,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87405		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

38	Q3345-07MS	TANK-FARM-WATER	MS	10/15/25 17:45		Jaswal	OK
39	Q3345-07MSD	TANK-FARM-WATER	MSD	10/15/25 17:49		Jaswal	OK
40	Q3345-07A	TANK-FARM-WATER	PS	10/15/25 17:53	0.1 ml of each M6008 AND M6017 IN 10ML SAMPLE.	Jaswal	OK
41	Q3150-06	MER Rd Con Ed	SAM	10/15/25 18:05		Jaswal	OK
42	CCV04	CCV04	CCV	10/15/25 18:10		Jaswal	OK
43	CCB04	CCB04	CCB	10/15/25 18:14		Jaswal	OK
44	Q3150-07	Mid Site Grid 5-7	SAM	10/15/25 18:18		Jaswal	OK
45	Q3345-01	TANK-FARM-SOIL	SAM	10/15/25 18:22		Jaswal	OK
46	Q3345-03	2ND-STREET-SOIL	SAM	10/15/25 18:26		Jaswal	OK
47	Q3345-03DUP	2ND-STREET-SOIL	DUP	10/15/25 18:30		Jaswal	OK
48	Q3345-03L	2ND-STREET-SOIL	SD	10/15/25 18:34		Jaswal	OK
49	Q3345-03MS	2ND-STREET-SOIL	MS	10/15/25 18:38		Jaswal	OK
50	Q3345-03MSD	2ND-STREET-SOIL	MSD	10/15/25 18:42		Jaswal	OK
51	Q3345-03A	2ND-STREET-SOIL	PS	10/15/25 18:46	0.1 ml of each M6008 AND M6017 IN 10ML SAMPLE.	Jaswal	OK
52	Q3345-05	M-R-SOIL	SAM	10/15/25 18:50		Jaswal	OK
53	Q3340-01	CONCRETE-1	SAM	10/15/25 18:54		Jaswal	OK
54	CCV05	CCV05	CCV	10/15/25 18:58		Jaswal	OK
55	CCB05	CCB05	CCB	10/15/25 19:02		Jaswal	OK
56	Q3340-02	CONCRETE-2	SAM	10/15/25 19:07		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137561

Review By	Janvi	Review On	10/17/2025 12:20:45 PM
Supervise By	jaswal	Supervise On	10/17/2025 1:18:42 PM

STD. NAME	STD REF.#
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032
ICV Standard	MP87404,MP87038
CCV Standard	MP87042
ICSA Standard	MP87040,MP87405
CRI Standard	MP87038
LCS Standard	
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049

57	Q3340-03	CONCRETE-3	SAM	10/15/25 19:11		Jaswal	OK
58	Q3341-01	C-1	SAM	10/15/25 19:15		Jaswal	OK
59	Q3341-02	C-2	SAM	10/15/25 19:19		Jaswal	OK
60	Q3341-03	C-3	SAM	10/15/25 19:23		Jaswal	OK
61	Q3342-01	CONC-ROLLOFF-WC	SAM	10/15/25 19:27		Jaswal	OK
62	Q3343-01	SOIL-ROLLOFF-WC	SAM	10/15/25 19:32		Jaswal	OK
63	Q3337-02	LAYDOWN-YARD-1	SAM	10/15/25 19:36		Jaswal	OK
64	Q3337-04	LAYDOWN-YARD-2	SAM	10/15/25 19:40		Jaswal	OK
65	CCV06	CCV06	CCV	10/15/25 19:44		Jaswal	OK
66	CCB06	CCB06	CCB	10/15/25 19:48		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137573

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/16/25 11:08		Jaswal	OK
2	S1	S1	CAL2	10/16/25 11:13		Jaswal	OK
3	S2	S2	CAL3	10/16/25 11:17		Jaswal	OK
4	S3	S3	CAL4	10/16/25 11:22		Jaswal	OK
5	S4	S4	CAL5	10/16/25 11:25		Jaswal	OK
6	S5	S5	CAL6	10/16/25 11:30		Jaswal	OK
7	ICV01	ICV01	ICV	10/16/25 11:34		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/16/25 11:58		Jaswal	OK
9	ICB01	ICB01	ICB	10/16/25 12:02		Jaswal	OK
10	CRI01	CRI01	CRDL	10/16/25 12:09		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/16/25 12:17		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/16/25 12:21		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/16/25 12:25		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/16/25 12:29		Jaswal	OK
15	CCV01	CCV01	CCV	10/16/25 12:38		Jaswal	OK
16	CCB01	CCB01	CCB	10/16/25 12:42		Jaswal	OK
17	LR-2	LR-2	HIGH STD	10/16/25 12:51		Jaswal	OK
18	LR-3	LR-3	HIGH STD	10/16/25 12:55		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137573

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

19	LR-1	LR-1	HIGH STD	10/16/25 13:00		Jaswal	OK
20	Q2893-01	LOD-MDL-SOIL-01-Q	SAM	10/16/25 13:04		Jaswal	OK
21	Q2893-16	LLOQ-SOIL-QT3-202	SAM	10/16/25 13:13		Jaswal	OK
22	Q2893-07	LOD-MDL-WATER-01	SAM	10/16/25 13:17		Jaswal	OK
23	Q2893-08	LOQ-WATER-02-QT3	SAM	10/16/25 13:22		Jaswal	OK
24	Q2893-17	LLOQ-WATER-QT3-2	SAM	10/16/25 13:26		Jaswal	OK
25	Q2893-10	LOD-MDL-WATER-04	SAM	10/16/25 13:30		Jaswal	OK
26	CCV02	CCV02	CCV	10/16/25 13:45		Jaswal	OK
27	CCB02	CCB02	CCB	10/16/25 13:49		Jaswal	OK
28	Q2893-02	LOQ-SOIL-02-QT3-20	SAM	10/16/25 13:59		Jaswal	OK
29	Q2893-11	LOQ-WATER-05-QT3	SAM	10/16/25 14:07		Jaswal	OK
30	PB169558BL	PB169558BL	MB	10/16/25 14:12		Jaswal	OK
31	PB169558BS	PB169558BS	LCS	10/16/25 14:16		Jaswal	OK
32	PB169557BL	PB169557BL	MB	10/16/25 14:20		Jaswal	OK
33	PB169557BS	PB169557BS	LCS	10/16/25 14:24		Jaswal	OK
34	PB169559BL	PB169559BL	MB	10/16/25 14:28		Jaswal	OK
35	PB169559BS	PB169559BS	LCS	10/16/25 14:40		Jaswal	OK
36	CCV03	CCV03	CCV	10/16/25 14:45		Jaswal	OK
37	CCB03	CCB03	CCB	10/16/25 14:49		Jaswal	OK
38	PB170107BL	PB170107BL	MB	10/16/25 14:55		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137573

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

39	PB170107BS	PB170107BS	LCS	10/16/25 14:59		Jaswal	OK
40	PB170112BL	PB170112BL	MB	10/16/25 15:03		Jaswal	OK
41	PB170112BS	PB170112BS	LCS	10/16/25 15:08		Jaswal	OK
42	PB170081TB	PB170081TB	MB	10/16/25 15:12	Cr recovery fail	Jaswal	Not Ok
43	Q3340-05	CONCRETE-1	SAM	10/16/25 15:16		Jaswal	OK
44	Q3340-06	CONCRETE-2	SAM	10/16/25 15:21		Jaswal	OK
45	Q3340-07	CONCRETE-3	SAM	10/16/25 15:25		Jaswal	OK
46	Q3342-02	CONC-ROLLOFF-WC	SAM	10/16/25 15:29		Jaswal	OK
47	Q3343-04	SOIL-ROLLOFF-WC-	SAM	10/16/25 15:34		Jaswal	OK
48	CCV04	CCV04	CCV	10/16/25 15:38		Jaswal	OK
49	CCB04	CCB04	CCB	10/16/25 15:43		Jaswal	OK
50	Q3345-02	TANK-FARM-SOIL	SAM	10/16/25 15:47		Jaswal	OK
51	Q3345-04	2ND-STREET-SOIL	SAM	10/16/25 15:52		Jaswal	OK
52	Q3345-06	M-R-SOIL	SAM	10/16/25 15:56		Jaswal	OK
53	Q3346-01	Zeltmann Power Plant	SAM	10/16/25 16:01		Jaswal	OK
54	Q3348-04	WCS-TP1	SAM	10/16/25 16:05		Jaswal	OK
55	Q3348-08	WCS-TP2	SAM	10/16/25 16:10		Jaswal	OK
56	Q3348-08DUP	WCS-TP2DUP	DUP	10/16/25 16:14		Jaswal	OK
57	Q3348-08L	WCS-TP2L	SD	10/16/25 16:18		Jaswal	OK
58	Q3348-08MS	WCS-TP2MS	MS	10/16/25 16:23		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137573

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

59	Q3348-08MSD	WCS-TP2MSD	MSD	10/16/25 16:27		Jaswal	OK
60	CCV05	CCV05	CCV	10/16/25 16:31		Jaswal	OK
61	CCB05	CCB05	CCB	10/16/25 16:35		Jaswal	OK
62	Q3356-01	AUD-25-0169-AUD-25	SAM	10/16/25 16:44		Jaswal	OK
63	Q3356-01DUP	AUD-25-0169-AUD-25	DUP	10/16/25 16:48		Jaswal	OK
64	Q3348-08A	WCS-TP2A	PS	10/16/25 16:53		Jaswal	OK
65	Q3356-01L	AUD-25-0169-AUD-25	SD	10/16/25 16:57		Jaswal	OK
66	Q3356-01MS	AUD-25-0169-AUD-25	MS	10/16/25 17:01		Jaswal	OK
67	Q3356-01MSD	AUD-25-0169-AUD-25	MSD	10/16/25 17:05		Jaswal	OK
68	PB170130BL	PB170130BL	MB	10/16/25 17:18		Jaswal	OK
69	PB170130BS	PB170130BS	LCS	10/16/25 17:22		Jaswal	OK
70	Q3355-01	EO-03-101525	SAM	10/16/25 17:26		Jaswal	OK
71	Q3356-01A	AUD-25-0169-AUD-25	PS	10/16/25 17:30		Jaswal	OK
72	CCV06	CCV06	CCV	10/16/25 17:49		Jaswal	OK
73	CCB06	CCB06	CCB	10/16/25 17:53		Jaswal	OK
74	PB170129BL	PB170129BL	MB	10/16/25 17:57		Jaswal	OK
75	Q3348-01	WCS-TP1	SAM	10/16/25 18:05		Jaswal	OK
76	Q3348-01DUP	WCS-TP1DUP	DUP	10/16/25 18:09		Jaswal	OK
77	Q3348-01L	WCS-TP1L	SD	10/16/25 18:13		Jaswal	OK
78	Q3348-01MS	WCS-TP1MS	MS	10/16/25 18:18		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137573

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

79	Q3348-01MSD	WCS-TP1MSD	MSD	10/16/25 18:22		Jaswal	OK
80	Q3348-01A	WCS-TP1A	PS	10/16/25 18:25		Jaswal	OK
81	Q3348-05	WCS-TP2	SAM	10/16/25 18:29		Jaswal	OK
82	PB170129BS	PB170129BS	LCS	10/16/25 18:34		Jaswal	OK
83	Q3346-01DL	Zeltmann Power Plant	SAM	10/16/25 18:38	Not use	Jaswal	Not Ok
84	CCV07	CCV07	CCV	10/16/25 18:42		Jaswal	OK
85	CCB07	CCB07	CCB	10/16/25 18:46		Jaswal	OK
86	Q3354-01	HT1403	SAM	10/16/25 18:51		Jaswal	OK
87	Q3352-01	TP-3	SAM	10/16/25 18:55		Jaswal	OK
88	PB170132BL	PB170132BL	MB	10/16/25 18:59		Jaswal	OK
89	PB170132BS	PB170132BS	LCS	10/16/25 19:03		Jaswal	OK
90	Q3349-01	PURS Station	SAM	10/16/25 19:07		Jaswal	OK
91	Q3352-04	TP-3	SAM	10/16/25 19:11		Jaswal	OK
92	Q3354-02	HT1403	SAM	10/16/25 19:16		Jaswal	OK
93	Q3357-02	DRUM-2	SAM	10/16/25 19:20		Jaswal	OK
94	Q3357-03	DRUM-3	SAM	10/16/25 19:24		Jaswal	OK
95	Q3357-04	DRUM-4	SAM	10/16/25 19:29		Jaswal	OK
96	CCV08	CCV08	CCV	10/16/25 19:33		Jaswal	OK
97	CCB08	CCB08	CCB	10/16/25 19:37		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137573

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

98	Q3357-04DUP	DRUM-4DUP	DUP	10/16/25 19:41	CCV FAIL FOR Al, Ba,Mg,Na	Jaswal	Not Ok
99	Q3357-04L	DRUM-4L	SD	10/16/25 19:46	CCV FAIL FOR Al, Ba,Mg,Na	Jaswal	Not Ok
100	Q3357-04MS	DRUM-4MS	MS	10/16/25 19:50	CCV FAIL FOR Al, Ba,Mg,Na	Jaswal	Not Ok
101	Q3357-04MSD	DRUM-4MSD	MSD	10/16/25 19:54	CCV FAIL FOR Al, Ba,Mg,Na	Jaswal	Not Ok
102	Q3357-04A	DRUM-4A	PS	10/16/25 19:58	CCV FAIL FOR Al, Ba,Mg,Na	Jaswal	Not Ok
103	PB170116TB	PB170116TB	MB	10/16/25 20:03	CCV FAIL FOR Al, Ba,Mg,Na	Jaswal	Not Ok
104	CCV09	CCV09	CCV	10/16/25 20:07	CCV FAIL FOR Al, Ba,Mg,Na	Jaswal	OK
105	CCB09	CCB09	CCB	10/16/25 20:11		Jaswal	OK