



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB71	Mercury	0.076	+/-0.2	U	0.20	CV	10/15/2025	09:08	LB137528

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB29	Mercury	0.076	+/-0.2	U	0.20	CV	10/15/2025	09:12	LB137528
CCB30	Mercury	0.076	+/-0.2	U	0.20	CV	10/15/2025	09:44	LB137528
CCB31	Mercury	0.076	+/-0.2	U	0.20	CV	10/15/2025	10:23	LB137528
CCB32	Mercury	0.076	+/-0.2	U	0.20	CV	10/15/2025	10:32	LB137528

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	10/15/2025	14:24	LB137561
	Antimony	6.76	+/-25	U	50.0	P	10/15/2025	14:24	LB137561
	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	14:24	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	14:24	LB137561
	Beryllium	0.56	+/-3	U	6.00	P	10/15/2025	14:24	LB137561
	Boron	15.7	+/-50	U	100	P	10/15/2025	14:24	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	14:24	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	14:24	LB137561
	Cobalt	2.26	+/-15	U	30.0	P	10/15/2025	14:24	LB137561
	Copper	4.60	+/-10	U	20.0	P	10/15/2025	14:24	LB137561
	Iron	23.4	+/-50	U	100	P	10/15/2025	14:24	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	14:24	LB137561
	Manganese	5.94	+/-10	U	20.0	P	10/15/2025	14:24	LB137561
	Molybdenum	18.7	+/-100	U	200	P	10/15/2025	14:24	LB137561
	Nickel	3.06	+/-20	U	40.0	P	10/15/2025	14:24	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	14:24	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	14:24	LB137561
	Thallium	4.38	+/-20	U	40.0	P	10/15/2025	14:24	LB137561
	Vanadium	6.26	+/-20	U	40.0	P	10/15/2025	14:24	LB137561
	Zinc	16.7	+/-20	U	40.0	P	10/15/2025	14:24	LB137561

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	10/15/2025	15:13	LB137561
	Antimony	6.76	+/-25	U	50.0	P	10/15/2025	15:13	LB137561
	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	15:13	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	15:13	LB137561
	Beryllium	0.56	+/-3	U	6.00	P	10/15/2025	15:13	LB137561
	Boron	15.7	+/-50	U	100	P	10/15/2025	15:13	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	15:13	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	15:13	LB137561
	Cobalt	2.26	+/-15	U	30.0	P	10/15/2025	15:13	LB137561
	Copper	4.60	+/-10	U	20.0	P	10/15/2025	15:13	LB137561
	Iron	23.4	+/-50	U	100	P	10/15/2025	15:13	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	15:13	LB137561
	Manganese	5.94	+/-10	U	20.0	P	10/15/2025	15:13	LB137561
	Molybdenum	18.7	+/-100	U	200	P	10/15/2025	15:13	LB137561
	Nickel	3.06	+/-20	U	40.0	P	10/15/2025	15:13	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	15:13	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	15:13	LB137561
	Thallium	4.38	+/-20	U	40.0	P	10/15/2025	15:13	LB137561
	Vanadium	6.26	+/-20	U	40.0	P	10/15/2025	15:13	LB137561
	Zinc	16.7	+/-20	U	40.0	P	10/15/2025	15:13	LB137561
CCB02	Aluminum	11.3	+/-50	U	100	P	10/15/2025	16:32	LB137561
	Antimony	6.76	+/-25	U	50.0	P	10/15/2025	16:32	LB137561
	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	16:32	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	16:32	LB137561
	Beryllium	0.56	+/-3	U	6.00	P	10/15/2025	16:32	LB137561
	Boron	15.7	+/-50	U	100	P	10/15/2025	16:32	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	16:32	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	16:32	LB137561
	Cobalt	2.26	+/-15	U	30.0	P	10/15/2025	16:32	LB137561
	Copper	4.60	+/-10	U	20.0	P	10/15/2025	16:32	LB137561
	Iron	23.4	+/-50	U	100	P	10/15/2025	16:32	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	16:32	LB137561
	Manganese	5.94	+/-10	U	20.0	P	10/15/2025	16:32	LB137561
	Molybdenum	18.7	+/-100	U	200	P	10/15/2025	16:32	LB137561
	Nickel	3.06	+/-20	U	40.0	P	10/15/2025	16:32	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	16:32	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	16:32	LB137561
	Thallium	4.38	+/-20	U	40.0	P	10/15/2025	16:32	LB137561
	Vanadium	6.26	+/-20	U	40.0	P	10/15/2025	16:32	LB137561

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Zinc	16.7	+/-20	U	40.0	P	10/15/2025	16:32	LB137561
CCB03	Aluminum	11.3	+/-50	U	100	P	10/15/2025	17:25	LB137561
	Antimony	6.76	+/-25	U	50.0	P	10/15/2025	17:25	LB137561
	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	17:25	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	17:25	LB137561
	Beryllium	0.56	+/-3	U	6.00	P	10/15/2025	17:25	LB137561
	Boron	15.7	+/-50	U	100	P	10/15/2025	17:25	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	17:25	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	17:25	LB137561
	Cobalt	2.26	+/-15	U	30.0	P	10/15/2025	17:25	LB137561
	Copper	4.60	+/-10	U	20.0	P	10/15/2025	17:25	LB137561
	Iron	23.4	+/-50	U	100	P	10/15/2025	17:25	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	17:25	LB137561
	Manganese	5.94	+/-10	U	20.0	P	10/15/2025	17:25	LB137561
	Molybdenum	18.7	+/-100	U	200	P	10/15/2025	17:25	LB137561
	Nickel	3.06	+/-20	U	40.0	P	10/15/2025	17:25	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	17:25	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	17:25	LB137561
	Thallium	4.38	+/-20	U	40.0	P	10/15/2025	17:25	LB137561
	Vanadium	6.26	+/-20	U	40.0	P	10/15/2025	17:25	LB137561
	Zinc	16.7	+/-20	U	40.0	P	10/15/2025	17:25	LB137561
CCB04	Aluminum	11.3	+/-50	U	100	P	10/15/2025	18:14	LB137561
	Antimony	6.76	+/-25	U	50.0	P	10/15/2025	18:14	LB137561
	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	18:14	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	18:14	LB137561
	Beryllium	0.56	+/-3	U	6.00	P	10/15/2025	18:14	LB137561
	Boron	15.7	+/-50	U	100	P	10/15/2025	18:14	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	18:14	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	18:14	LB137561
	Cobalt	2.26	+/-15	U	30.0	P	10/15/2025	18:14	LB137561
	Copper	4.60	+/-10	U	20.0	P	10/15/2025	18:14	LB137561
	Iron	23.4	+/-50	U	100	P	10/15/2025	18:14	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	18:14	LB137561
	Manganese	5.94	+/-10	U	20.0	P	10/15/2025	18:14	LB137561
	Molybdenum	18.7	+/-100	U	200	P	10/15/2025	18:14	LB137561
	Nickel	3.06	+/-20	U	40.0	P	10/15/2025	18:14	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	18:14	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	18:14	LB137561
	Thallium	4.38	+/-20	U	40.0	P	10/15/2025	18:14	LB137561

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Vanadium	6.26	+/-20	U	40.0	P	10/15/2025	18:14	LB137561
	Zinc	16.7	+/-20	U	40.0	P	10/15/2025	18:14	LB137561
CCB05	Aluminum	11.3	+/-50	U	100	P	10/15/2025	19:02	LB137561
	Antimony	6.76	+/-25	U	50.0	P	10/15/2025	19:02	LB137561
	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	19:02	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	19:02	LB137561
	Beryllium	0.56	+/-3	U	6.00	P	10/15/2025	19:02	LB137561
	Boron	15.7	+/-50	U	100	P	10/15/2025	19:02	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	19:02	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	19:02	LB137561
	Cobalt	2.26	+/-15	U	30.0	P	10/15/2025	19:02	LB137561
	Copper	4.60	+/-10	U	20.0	P	10/15/2025	19:02	LB137561
	Iron	23.4	+/-50	U	100	P	10/15/2025	19:02	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	19:02	LB137561
	Manganese	5.94	+/-10	U	20.0	P	10/15/2025	19:02	LB137561
	Molybdenum	18.7	+/-100	U	200	P	10/15/2025	19:02	LB137561
	Nickel	3.06	+/-20	U	40.0	P	10/15/2025	19:02	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	19:02	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	19:02	LB137561
	Thallium	4.38	+/-20	U	40.0	P	10/15/2025	19:02	LB137561
	Vanadium	6.26	+/-20	U	40.0	P	10/15/2025	19:02	LB137561
	Zinc	16.7	+/-20	U	40.0	P	10/15/2025	19:02	LB137561
CCB06	Aluminum	11.3	+/-50	U	100	P	10/15/2025	19:48	LB137561
	Antimony	6.76	+/-25	U	50.0	P	10/15/2025	19:48	LB137561
	Arsenic	5.12	+/-10	U	20.0	P	10/15/2025	19:48	LB137561
	Barium	14.6	+/-50	U	100	P	10/15/2025	19:48	LB137561
	Beryllium	0.56	+/-3	U	6.00	P	10/15/2025	19:48	LB137561
	Boron	15.7	+/-50	U	100	P	10/15/2025	19:48	LB137561
	Cadmium	0.50	+/-3	U	6.00	P	10/15/2025	19:48	LB137561
	Chromium	2.12	+/-5	U	10.0	P	10/15/2025	19:48	LB137561
	Cobalt	2.26	+/-15	U	30.0	P	10/15/2025	19:48	LB137561
	Copper	4.60	+/-10	U	20.0	P	10/15/2025	19:48	LB137561
	Iron	23.4	+/-50	U	100	P	10/15/2025	19:48	LB137561
	Lead	2.30	+/-6	U	12.0	P	10/15/2025	19:48	LB137561
	Manganese	5.94	+/-10	U	20.0	P	10/15/2025	19:48	LB137561
	Molybdenum	18.7	+/-100	U	200	P	10/15/2025	19:48	LB137561
	Nickel	3.06	+/-20	U	40.0	P	10/15/2025	19:48	LB137561
	Selenium	9.64	+/-10	U	20.0	P	10/15/2025	19:48	LB137561
	Silver	1.62	+/-5	U	10.0	P	10/15/2025	19:48	LB137561

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Thallium	4.38	+/-20	U	40.0	P	10/15/2025	19:48	LB137561
	Vanadium	6.26	+/-20	U	40.0	P	10/15/2025	19:48	LB137561
	Zinc	16.7	+/-20	U	40.0	P	10/15/2025	19:48	LB137561

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	10/16/2025	12:02	LB137573
	Antimony	6.76	+/-25	U	50.0	P	10/16/2025	12:02	LB137573
	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	12:02	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	12:02	LB137573
	Beryllium	0.56	+/-3	U	6.00	P	10/16/2025	12:02	LB137573
	Boron	15.7	+/-50	U	100	P	10/16/2025	12:02	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	12:02	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	12:02	LB137573
	Cobalt	2.26	+/-15	U	30.0	P	10/16/2025	12:02	LB137573
	Copper	4.60	+/-10	U	20.0	P	10/16/2025	12:02	LB137573
	Iron	23.4	+/-50	U	100	P	10/16/2025	12:02	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	12:02	LB137573
	Manganese	5.94	+/-10	U	20.0	P	10/16/2025	12:02	LB137573
	Molybdenum	18.7	+/-100	U	200	P	10/16/2025	12:02	LB137573
	Nickel	3.06	+/-20	U	40.0	P	10/16/2025	12:02	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	12:02	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	12:02	LB137573
	Thallium	4.38	+/-20	U	40.0	P	10/16/2025	12:02	LB137573
	Vanadium	6.26	+/-20	U	40.0	P	10/16/2025	12:02	LB137573
	Zinc	16.7	+/-20	U	40.0	P	10/16/2025	12:02	LB137573

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	12.2	+/-50	J	100	P	10/16/2025	12:42	LB137573
	Antimony	6.76	+/-25	U	50.0	P	10/16/2025	12:42	LB137573
	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	12:42	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	12:42	LB137573
	Beryllium	0.56	+/-3	U	6.00	P	10/16/2025	12:42	LB137573
	Boron	15.7	+/-50	U	100	P	10/16/2025	12:42	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	12:42	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	12:42	LB137573
	Cobalt	2.26	+/-15	U	30.0	P	10/16/2025	12:42	LB137573
	Copper	4.60	+/-10	U	20.0	P	10/16/2025	12:42	LB137573
	Iron	23.4	+/-50	U	100	P	10/16/2025	12:42	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	12:42	LB137573
	Manganese	5.94	+/-10	U	20.0	P	10/16/2025	12:42	LB137573
	Molybdenum	18.7	+/-100	U	200	P	10/16/2025	12:42	LB137573
	Nickel	3.06	+/-20	U	40.0	P	10/16/2025	12:42	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	12:42	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	12:42	LB137573
	Thallium	4.38	+/-20	U	40.0	P	10/16/2025	12:42	LB137573
	Vanadium	6.26	+/-20	U	40.0	P	10/16/2025	12:42	LB137573
	Zinc	16.7	+/-20	U	40.0	P	10/16/2025	12:42	LB137573
CCB02	Aluminum	31.3	+/-50	J	100	P	10/16/2025	13:49	LB137573
	Antimony	6.76	+/-25	U	50.0	P	10/16/2025	13:49	LB137573
	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	13:49	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	13:49	LB137573
	Beryllium	0.56	+/-3	U	6.00	P	10/16/2025	13:49	LB137573
	Boron	15.7	+/-50	U	100	P	10/16/2025	13:49	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	13:49	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	13:49	LB137573
	Cobalt	2.26	+/-15	U	30.0	P	10/16/2025	13:49	LB137573
	Copper	4.60	+/-10	U	20.0	P	10/16/2025	13:49	LB137573
	Iron	23.4	+/-50	U	100	P	10/16/2025	13:49	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	13:49	LB137573
	Manganese	5.94	+/-10	U	20.0	P	10/16/2025	13:49	LB137573
	Molybdenum	18.7	+/-100	U	200	P	10/16/2025	13:49	LB137573
	Nickel	3.06	+/-20	U	40.0	P	10/16/2025	13:49	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	13:49	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	13:49	LB137573
	Thallium	4.38	+/-20	U	40.0	P	10/16/2025	13:49	LB137573
	Vanadium	6.26	+/-20	U	40.0	P	10/16/2025	13:49	LB137573

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Zinc	16.7	+/-20	U	40.0	P	10/16/2025	13:49	LB137573
CCB03	Aluminum	19.5	+/-50	J	100	P	10/16/2025	14:49	LB137573
	Antimony	6.76	+/-25	U	50.0	P	10/16/2025	14:49	LB137573
	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	14:49	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	14:49	LB137573
	Beryllium	0.56	+/-3	U	6.00	P	10/16/2025	14:49	LB137573
	Boron	15.7	+/-50	U	100	P	10/16/2025	14:49	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	14:49	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	14:49	LB137573
	Cobalt	2.26	+/-15	U	30.0	P	10/16/2025	14:49	LB137573
	Copper	4.60	+/-10	U	20.0	P	10/16/2025	14:49	LB137573
	Iron	23.4	+/-50	U	100	P	10/16/2025	14:49	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	14:49	LB137573
	Manganese	5.94	+/-10	U	20.0	P	10/16/2025	14:49	LB137573
	Molybdenum	18.7	+/-100	U	200	P	10/16/2025	14:49	LB137573
	Nickel	3.06	+/-20	U	40.0	P	10/16/2025	14:49	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	14:49	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	14:49	LB137573
	Thallium	4.38	+/-20	U	40.0	P	10/16/2025	14:49	LB137573
	Vanadium	6.26	+/-20	U	40.0	P	10/16/2025	14:49	LB137573
	Zinc	16.7	+/-20	U	40.0	P	10/16/2025	14:49	LB137573
CCB04	Aluminum	32.4	+/-50	J	100	P	10/16/2025	15:43	LB137573
	Antimony	6.76	+/-25	U	50.0	P	10/16/2025	15:43	LB137573
	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	15:43	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	15:43	LB137573
	Beryllium	0.56	+/-3	U	6.00	P	10/16/2025	15:43	LB137573
	Boron	15.7	+/-50	U	100	P	10/16/2025	15:43	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	15:43	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	15:43	LB137573
	Cobalt	2.26	+/-15	U	30.0	P	10/16/2025	15:43	LB137573
	Copper	4.60	+/-10	U	20.0	P	10/16/2025	15:43	LB137573
	Iron	23.4	+/-50	U	100	P	10/16/2025	15:43	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	15:43	LB137573
	Manganese	5.94	+/-10	U	20.0	P	10/16/2025	15:43	LB137573
	Molybdenum	18.7	+/-100	U	200	P	10/16/2025	15:43	LB137573
	Nickel	3.06	+/-20	U	40.0	P	10/16/2025	15:43	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	15:43	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	15:43	LB137573
	Thallium	15.8	+/-20	J	40.0	P	10/16/2025	15:43	LB137573

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Vanadium	6.26	+/-20	U	40.0	P	10/16/2025	15:43	LB137573
	Zinc	16.7	+/-20	U	40.0	P	10/16/2025	15:43	LB137573
CCB05	Aluminum	13.5	+/-50	J	100	P	10/16/2025	16:35	LB137573
	Antimony	6.76	+/-25	U	50.0	P	10/16/2025	16:35	LB137573
	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	16:35	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	16:35	LB137573
	Beryllium	0.56	+/-3	U	6.00	P	10/16/2025	16:35	LB137573
	Boron	15.7	+/-50	U	100	P	10/16/2025	16:35	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	16:35	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	16:35	LB137573
	Cobalt	2.26	+/-15	U	30.0	P	10/16/2025	16:35	LB137573
	Copper	4.60	+/-10	U	20.0	P	10/16/2025	16:35	LB137573
	Iron	23.4	+/-50	U	100	P	10/16/2025	16:35	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	16:35	LB137573
	Manganese	5.94	+/-10	U	20.0	P	10/16/2025	16:35	LB137573
	Molybdenum	18.7	+/-100	U	200	P	10/16/2025	16:35	LB137573
	Nickel	3.06	+/-20	U	40.0	P	10/16/2025	16:35	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	16:35	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	16:35	LB137573
	Thallium	14.1	+/-20	J	40.0	P	10/16/2025	16:35	LB137573
	Vanadium	6.26	+/-20	U	40.0	P	10/16/2025	16:35	LB137573
	Zinc	16.7	+/-20	U	40.0	P	10/16/2025	16:35	LB137573
CCB06	Aluminum	14.7	+/-50	J	100	P	10/16/2025	17:53	LB137573
	Antimony	6.76	+/-25	U	50.0	P	10/16/2025	17:53	LB137573
	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	17:53	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	17:53	LB137573
	Beryllium	0.56	+/-3	U	6.00	P	10/16/2025	17:53	LB137573
	Boron	15.7	+/-50	U	100	P	10/16/2025	17:53	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	17:53	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	17:53	LB137573
	Cobalt	2.26	+/-15	U	30.0	P	10/16/2025	17:53	LB137573
	Copper	4.60	+/-10	U	20.0	P	10/16/2025	17:53	LB137573
	Iron	28.7	+/-50	J	100	P	10/16/2025	17:53	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	17:53	LB137573
	Manganese	5.94	+/-10	U	20.0	P	10/16/2025	17:53	LB137573
	Molybdenum	18.7	+/-100	U	200	P	10/16/2025	17:53	LB137573
	Nickel	3.06	+/-20	U	40.0	P	10/16/2025	17:53	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	17:53	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	17:53	LB137573

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Thallium	4.38	+/-20	U	40.0	P	10/16/2025	17:53	LB137573
	Vanadium	6.26	+/-20	U	40.0	P	10/16/2025	17:53	LB137573
	Zinc	16.7	+/-20	U	40.0	P	10/16/2025	17:53	LB137573
CCB07	Aluminum	11.3	+/-50	U	100	P	10/16/2025	18:46	LB137573
	Antimony	6.76	+/-25	U	50.0	P	10/16/2025	18:46	LB137573
	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	18:46	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	18:46	LB137573
	Beryllium	0.56	+/-3	U	6.00	P	10/16/2025	18:46	LB137573
	Boron	15.7	+/-50	U	100	P	10/16/2025	18:46	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	18:46	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	18:46	LB137573
	Cobalt	2.26	+/-15	U	30.0	P	10/16/2025	18:46	LB137573
	Copper	4.60	+/-10	U	20.0	P	10/16/2025	18:46	LB137573
	Iron	23.4	+/-50	U	100	P	10/16/2025	18:46	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	18:46	LB137573
	Manganese	5.94	+/-10	U	20.0	P	10/16/2025	18:46	LB137573
	Molybdenum	18.7	+/-100	U	200	P	10/16/2025	18:46	LB137573
	Nickel	3.06	+/-20	U	40.0	P	10/16/2025	18:46	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	18:46	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	18:46	LB137573
	Thallium	4.38	+/-20	U	40.0	P	10/16/2025	18:46	LB137573
	Vanadium	6.26	+/-20	U	40.0	P	10/16/2025	18:46	LB137573
	Zinc	16.7	+/-20	U	40.0	P	10/16/2025	18:46	LB137573
CCB08	Aluminum	11.3	+/-50	U	100	P	10/16/2025	19:37	LB137573
	Antimony	6.76	+/-25	U	50.0	P	10/16/2025	19:37	LB137573
	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	19:37	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	19:37	LB137573
	Beryllium	0.56	+/-3	U	6.00	P	10/16/2025	19:37	LB137573
	Boron	15.7	+/-50	U	100	P	10/16/2025	19:37	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	19:37	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	19:37	LB137573
	Cobalt	2.26	+/-15	U	30.0	P	10/16/2025	19:37	LB137573
	Copper	4.60	+/-10	U	20.0	P	10/16/2025	19:37	LB137573
	Iron	23.4	+/-50	U	100	P	10/16/2025	19:37	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	19:37	LB137573
	Manganese	5.94	+/-10	U	20.0	P	10/16/2025	19:37	LB137573
	Molybdenum	18.7	+/-100	U	200	P	10/16/2025	19:37	LB137573
	Nickel	3.06	+/-20	U	40.0	P	10/16/2025	19:37	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	19:37	LB137573

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Silver	1.62	+/-5	U	10.0	P	10/16/2025	19:37	LB137573
	Thallium	4.38	+/-20	U	40.0	P	10/16/2025	19:37	LB137573
	Vanadium	6.26	+/-20	U	40.0	P	10/16/2025	19:37	LB137573
	Zinc	16.7	+/-20	U	40.0	P	10/16/2025	19:37	LB137573
CCB09	Aluminum	11.3	+/-50	U	100	P	10/16/2025	20:11	LB137573
	Antimony	6.76	+/-25	U	50.0	P	10/16/2025	20:11	LB137573
	Arsenic	5.12	+/-10	U	20.0	P	10/16/2025	20:11	LB137573
	Barium	14.6	+/-50	U	100	P	10/16/2025	20:11	LB137573
	Beryllium	0.56	+/-3	U	6.00	P	10/16/2025	20:11	LB137573
	Boron	15.7	+/-50	U	100	P	10/16/2025	20:11	LB137573
	Cadmium	0.50	+/-3	U	6.00	P	10/16/2025	20:11	LB137573
	Chromium	2.12	+/-5	U	10.0	P	10/16/2025	20:11	LB137573
	Cobalt	2.26	+/-15	U	30.0	P	10/16/2025	20:11	LB137573
	Copper	4.60	+/-10	U	20.0	P	10/16/2025	20:11	LB137573
	Iron	23.4	+/-50	U	100	P	10/16/2025	20:11	LB137573
	Lead	2.30	+/-6	U	12.0	P	10/16/2025	20:11	LB137573
	Manganese	5.94	+/-10	U	20.0	P	10/16/2025	20:11	LB137573
	Molybdenum	18.7	+/-100	U	200	P	10/16/2025	20:11	LB137573
	Nickel	3.06	+/-20	U	40.0	P	10/16/2025	20:11	LB137573
	Selenium	9.64	+/-10	U	20.0	P	10/16/2025	20:11	LB137573
	Silver	1.62	+/-5	U	10.0	P	10/16/2025	20:11	LB137573
	Thallium	4.38	+/-20	U	40.0	P	10/16/2025	20:11	LB137573
	Vanadium	6.26	+/-20	U	40.0	P	10/16/2025	20:11	LB137573
	Zinc	16.7	+/-20	U	40.0	P	10/16/2025	20:11	LB137573