

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170097BS Mercury	mg/Kg	0.28	0.27		96	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kleinfelder

SDG No.: Q3341

Contract: POWE02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170107BS							
Aluminum	mg/Kg	100	90.5		90	80 - 120	P
Antimony	mg/Kg	40.0	37.9		95	80 - 120	P
Arsenic	mg/Kg	40.0	36.4		91	80 - 120	P
Barium	mg/Kg	10.0	8.65		86	80 - 120	P
Beryllium	mg/Kg	10.0	9.39		94	80 - 120	P
Boron	mg/Kg	15.0	13.7		91	80 - 120	P
Cadmium	mg/Kg	10.0	8.61		86	80 - 120	P
Chromium	mg/Kg	20.0	18.7		94	80 - 120	P
Cobalt	mg/Kg	10.0	8.84		88	80 - 120	P
Copper	mg/Kg	15.0	14.4		96	80 - 120	P
Iron	mg/Kg	150	136		91	80 - 120	P
Lead	mg/Kg	50.0	43.4		87	80 - 120	P
Manganese	mg/Kg	10.0	9.09		91	80 - 120	P
Molybdenum	mg/Kg	20.0	19.0		95	80 - 120	P
Nickel	mg/Kg	25.0	22.2		89	80 - 120	P
Selenium	mg/Kg	100	93.8		94	80 - 120	P
Silver	mg/Kg	3.8	3.49		92	80 - 120	P
Thallium	mg/Kg	100	106		106	80 - 120	P
Vanadium	mg/Kg	15.0	13.5		90	80 - 120	P
Zinc	mg/Kg	10.0	9.63		96	80 - 120	P