

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: POWE02
SDG NO.: Q3341

Calibration Date(s): 10/09/2025 10/09/2025
Calibration Times: 14:14 22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO114313.D	CF 750 =	PO114314.D			
CF 500 =		PO114315.D	CF 250 =	PO114316.D	CF 050 =	PO114317.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	174192886	179895124	183856868	183322864	180520760	180357700	2
Aroclor-1016-2	(2)	293939915	305561101	305220730	315140900	301624180	304297365	3
Aroclor-1016-3	(3)	165633600	172668037	174260432	182505220	189193940	176852246	5
Aroclor-1016-4	(4)	134425999	138097755	140122574	142368188	132758060	137554515	3
Aroclor-1016-5	(5)	123446822	127769624	128908938	130194584	121008000	126265594	3
Aroclor-1260-1	(1)	213791672	221280347	222839010	218586200	213158120	217931070	2
Aroclor-1260-2	(2)	302622462	312056085	309594538	307167896	311081240	308504444	1
Aroclor-1260-3	(3)	259467311	268042008	265717042	260457100	228668800	256470452	6
Aroclor-1260-4	(4)	230199629	235131995	233767838	232769056	232729060	232919516	1
Aroclor-1260-5	(5)	586428065	605840996	593634282	581137220	533761820	580160477	5
Decachlorobiphenyl		3664039320	3792684880	3728036680	3664938040	3355866400	3641113064	5
Tetrachloro-m-xylene		5506498360	5677498373	5642459220	5552412920	4766000200	5428973815	7
Decachlorobiphenyl		3803036960	3909619067	3861003240	3833959680	3524921000	3786507989	4
Tetrachloro-m-xylene		5804343740	5928892613	5893981820	5915994040	5454317000	5799505843	3
Decachlorobiphenyl		3816291730	3937073800	3957474960	3763925080	3735520600	3842057234	3
Tetrachloro-m-xylene		5779379270	6044791093	6117390360	5832290520	5648377200	5884445689	3
Aroclor-1254-1	(1)	171902953	180224452	182326136	179059532	158754700	174453555	6
Aroclor-1254-2	(2)	248603985	257164111	266428592	263066468	238270000	254706631	4
Aroclor-1254-3	(3)	278966197	288101712	293337044	293741552	290556100	288940521	2
Aroclor-1254-4	(4)	232417367	241793461	248310406	242500032	218210100	236646273	5
Aroclor-1254-5	(5)	221022379	230048296	231810568	224607672	189890120	219475807	8
Decachlorobiphenyl		3792373920	3896541893	3936938560	3860958960	3451332800	3787629227	5
Tetrachloro-m-xylene		5822824890	5988302693	6053484060	5960886280	5298842000	5824867985	5
Decachlorobiphenyl		7124253780	7355062013	7282856320	7142144040	6358111400	7052485511	6
Tetrachloro-m-xylene		6131677880	6338649440	6321767300	6220727560	5845855000	6171735436	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: POWE02
SDG NO.: Q3341

Calibration Date(s): 10/09/2025 10/09/2025
Calibration Times: 14:14 22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO114313.D	CF 750 =	PO114314.D			
CF 500 =		PO114315.D	CF 250 =	PO114316.D	CF 050 =	PO114317.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	158257841	168187716	172362186	178585052	164529320	168384423	5
Aroclor-1016-2	(2)	229568804	240612359	245734900	252904664	241358100	242035765	4
Aroclor-1016-3	(3)	119655603	124240307	128221016	131086968	117564320	124153643	5
Aroclor-1016-4	(4)	92316708	96506141	100242646	102698056	90863720	96525454	5
Aroclor-1016-5	(5)	119349433	124629577	128669076	134240076	132389300	127855492	5
Aroclor-1260-1	(1)	215558470	231498535	235626690	245673392	249059400	235483297	6
Aroclor-1260-2	(2)	315603721	330098711	337167224	350489224	338584660	334388708	4
Aroclor-1260-3	(3)	256430713	268520155	278444642	296438324	374872640	294941295	16
Aroclor-1260-4	(4)	211317075	223266731	226148960	232910768	220902880	222909283	4
Aroclor-1260-5	(5)	594775480	624015233	623481606	632255524	590282220	612962013	3
Decachlorobiphenyl		3170001190	3348740240	3399801360	3490868680	3256491400	3333180574	4
Tetrachloro-m-xylene		3928229090	4070663293	4096822280	4116538520	3799017800	4002254197	3
Decachlorobiphenyl		3517109590	3681723480	3757541820	3835681800	3777005200	3713812378	3
Tetrachloro-m-xylene		4432077070	4412224160	4379772960	4471865640	4248743200	4388936606	2
Decachlorobiphenyl		3661201490	3833278560	3951800380	3911090520	3937109200	3858896030	3
Tetrachloro-m-xylene		4430508460	4634013587	4677547100	4537709360	4261034200	4508162541	4
Aroclor-1254-1	(1)	247222467	259048601	270361008	279309448	272529160	265694137	5
Aroclor-1254-2	(2)	219192417	228941475	239248508	249551080	254226320	238231960	6
Aroclor-1254-3	(3)	340127494	353950264	366242670	368614488	343101920	354407367	4
Aroclor-1254-4	(4)	246402938	256874435	266498340	267498992	253508600	258156661	3
Aroclor-1254-5	(5)	309302184	322642697	332806976	330987116	305278780	320203551	4
Decachlorobiphenyl		3673134130	3816212200	3925129200	3949127480	3714385800	3815597762	3
Tetrachloro-m-xylene		4426752370	4520704027	4624441060	4560126680	4117729800	4449950787	4
Decachlorobiphenyl		6770675710	7072875573	7158178740	7244610400	7053861200	7060040325	3
Tetrachloro-m-xylene		4643395640	4793601640	4792629120	4749683120	4526037200	4701069344	2