

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3341

Continuing Calib Date: 10/15/2025

Initial Calibration Date(s): 10/09/2025 10/09/2025

Continuing Calib Time: 16:45

Initial Calibration Time(s): 14:14 22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	4.36	4.36	4.26	4.46	0.00
Decachlorobiphenyl	9.94	9.93	9.83	10.03	-0.01

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** POWE02

**Lab Code:** ACE

**SDG NO.:** Q3341

**Continuing Calib Date:** 10/15/2025

**Initial Calibration Date(s):** 10/09/2025 10/09/2025

**Continuing Calib Time:** 16:45

**Initial Calibration Time(s):** 14:14 22:29

**GC Column:** ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM   TO		DIFF RT
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.67	8.67	8.57	8.77	0.00