

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3341

Continuing Calib Date: 10/15/2025

Initial Calibration Date(s): 09/23/2025 09/23/2025

Continuing Calib Time: 16:18

Initial Calibration Time(s): 08:45 17:14

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Tetrachloro-m-xylene	4.65	4.66	4.56	4.76	0.01
Decachlorobiphenyl	10.43	10.44	10.34	10.54	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3341

Continuing Calib Date: 10/15/2025

Initial Calibration Date(s): 09/23/2025 09/23/2025

Continuing Calib Time: 16:18

Initial Calibration Time(s): 08:45 17:14

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	3.79	3.80	3.70	3.90	0.01
Decachlorobiphenyl	8.79	8.80	8.70	8.90	0.01