

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114399.D **Time Analyzed:** 09:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.494	5.392	5.592	551.310	500.000	10.3
Aroclor-1016-2	5.515	5.414	5.614	549.040	500.000	9.8
Aroclor-1016-3	5.577	5.476	5.676	550.040	500.000	10.0
Aroclor-1016-4	5.673	5.571	5.771	581.160	500.000	16.2
Aroclor-1016-5	5.964	5.863	6.063	568.820	500.000	13.8
Aroclor-1260-1	7.077	6.975	7.175	512.110	500.000	2.4
Aroclor-1260-2	7.331	7.229	7.429	506.580	500.000	1.3
Aroclor-1260-3	7.687	7.585	7.785	515.490	500.000	3.1
Aroclor-1260-4	7.909	7.807	8.007	496.480	500.000	-0.7
Aroclor-1260-5	8.215	8.113	8.313	487.800	500.000	-2.4
Decachlorobiphenyl	9.938	9.834	10.034	50.550	50.000	1.1
Tetrachloro-m-xylene	4.356	4.255	4.455	56.960	50.000	13.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114399.D **Time Analyzed:** 09:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.742	4.643	4.843	550.910	500.000	10.2
Aroclor-1016-2	4.761	4.662	4.862	539.480	500.000	7.9
Aroclor-1016-3	4.936	4.836	5.036	562.380	500.000	12.5
Aroclor-1016-4	4.978	4.879	5.079	569.530	500.000	13.9
Aroclor-1016-5	5.190	5.092	5.292	556.930	500.000	11.4
Aroclor-1260-1	6.223	6.125	6.325	569.420	500.000	13.9
Aroclor-1260-2	6.411	6.313	6.513	558.200	500.000	11.6
Aroclor-1260-3	6.564	6.466	6.666	498.490	500.000	-0.3
Aroclor-1260-4	7.036	6.939	7.139	526.270	500.000	5.3
Aroclor-1260-5	7.276	7.179	7.379	495.490	500.000	-0.9
Decachlorobiphenyl	8.667	8.571	8.771	43.950	50.000	-12.1
Tetrachloro-m-xylene	3.660	3.560	3.760	54.880	50.000	9.8