

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114414.D **Time Analyzed:** 16:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.496	5.392	5.592	518.660	500.000	3.7
Aroclor-1016-2	5.517	5.414	5.614	507.260	500.000	1.5
Aroclor-1016-3	5.579	5.476	5.676	495.340	500.000	-0.9
Aroclor-1016-4	5.674	5.571	5.771	529.070	500.000	5.8
Aroclor-1016-5	5.966	5.863	6.063	524.990	500.000	5.0
Aroclor-1260-1	7.079	6.975	7.175	512.700	500.000	2.5
Aroclor-1260-2	7.333	7.229	7.429	491.540	500.000	-1.7
Aroclor-1260-3	7.690	7.585	7.785	496.280	500.000	-0.7
Aroclor-1260-4	7.912	7.807	8.007	488.760	500.000	-2.2
Aroclor-1260-5	8.218	8.113	8.313	481.920	500.000	-3.6
Decachlorobiphenyl	9.943	9.834	10.034	49.020	50.000	-2.0
Tetrachloro-m-xylene	4.357	4.255	4.455	54.300	50.000	8.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114414.D **Time Analyzed:** 16:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.743	4.643	4.843	530.650	500.000	6.1
Aroclor-1016-2	4.762	4.662	4.862	530.310	500.000	6.1
Aroclor-1016-3	4.937	4.836	5.036	551.950	500.000	10.4
Aroclor-1016-4	4.979	4.879	5.079	548.750	500.000	9.8
Aroclor-1016-5	5.192	5.092	5.292	533.500	500.000	6.7
Aroclor-1260-1	6.225	6.125	6.325	559.490	500.000	11.9
Aroclor-1260-2	6.413	6.313	6.513	554.670	500.000	10.9
Aroclor-1260-3	6.566	6.466	6.666	505.110	500.000	1.0
Aroclor-1260-4	7.038	6.939	7.139	536.640	500.000	7.3
Aroclor-1260-5	7.278	7.179	7.379	511.590	500.000	2.3
Decachlorobiphenyl	8.669	8.571	8.771	44.200	50.000	-11.6
Tetrachloro-m-xylene	3.661	3.560	3.760	53.500	50.000	7.0