

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075801.D **Time Analyzed:** 15:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.807	5.708	5.908	513.930	500.000	2.8
Aroclor-1016-2	5.829	5.729	5.929	519.260	500.000	3.9
Aroclor-1016-3	5.892	5.792	5.992	499.150	500.000	-0.2
Aroclor-1016-4	5.989	5.889	6.089	533.780	500.000	6.8
Aroclor-1016-5	6.281	6.181	6.381	534.710	500.000	6.9
Aroclor-1260-1	7.398	7.299	7.499	498.360	500.000	-0.3
Aroclor-1260-2	7.651	7.551	7.751	491.270	500.000	-1.7
Aroclor-1260-3	8.010	7.909	8.109	534.130	500.000	6.8
Aroclor-1260-4	8.238	8.137	8.337	549.030	500.000	9.8
Aroclor-1260-5	8.563	8.464	8.664	542.640	500.000	8.5
Decachlorobiphenyl	10.434	10.334	10.534	42.640	50.000	-14.7
Tetrachloro-m-xylene	4.655	4.555	4.755	47.020	50.000	-6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075801.D **Time Analyzed:** 15:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.889	4.793	4.993	515.970	500.000	3.2
Aroclor-1016-2	4.946	4.851	5.051	540.120	500.000	8.0
Aroclor-1016-3	5.066	4.971	5.171	490.520	500.000	-1.9
Aroclor-1016-4	5.106	5.011	5.211	526.890	500.000	5.4
Aroclor-1016-5	5.321	5.225	5.425	528.950	500.000	5.8
Aroclor-1260-1	6.349	6.255	6.455	478.440	500.000	-4.3
Aroclor-1260-2	6.536	6.441	6.641	493.970	500.000	-1.2
Aroclor-1260-3	6.689	6.595	6.795	503.510	500.000	0.7
Aroclor-1260-4	7.159	7.065	7.265	441.490	500.000	-11.7
Aroclor-1260-5	7.399	7.304	7.504	466.040	500.000	-6.8
Decachlorobiphenyl	8.795	8.703	8.903	47.140	50.000	-5.7
Tetrachloro-m-xylene	3.793	3.696	3.896	47.140	50.000	-5.7