

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3341

Continuing Calib Date: 10/15/2025

Initial Calibration Date(s): 09/18/2025 09/18/2025

Continuing Calib Time: 17:55

Initial Calibration Time(s): 10:30 11:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.07	9.07	8.97	9.17	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
Aldrin	5.27	5.27	5.17	5.37	0.01
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: POWE02

Lab Code: ACE

SDG NO.: Q3341

Continuing Calib Date: 10/15/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 17:55

Initial Calibration Time(s): 10:30

11:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
Aldrin	4.36	4.36	4.26	4.46	0.00
Dieldrin	5.50	5.51	5.41	5.61	0.01
4,4'-DDE	5.37	5.37	5.27	5.47	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
4,4'-DDT	6.17	6.18	6.08	6.28	0.01